

Continuing Calibration Report for SequenceID : FE092925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE056077.D

Aliphatic C12-C16	6293193.000	40.000	6.947	10.396	157329.825	153376.491	-2.578
Aliphatic C16-C21	10298086.000	60.000	10.397	13.773	171634.767	165879.737	-3.469
Aliphatic C21-C28	11678563.000	80.000	13.774	17.443	145982.038	148527.151	1.714
Aliphatic C28-C40	15804133.000	120.000	17.444	22.449	131701.108	136251.923	3.340
Aliphatic EPH	51783287.000	360.000	3.315	22.449	143842.464	146449.789	1.780
Aliphatic C9-C12	7709312.000	60.000	3.315	6.946	128488.533	140027.957	8.241