

Continuing Calibration Report for SequenceID : FF092925AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016469.D

Aromatic C10-C12	5084345.000	40.000	4.521	6.397	127108.625	143322.298	11.313
Aromatic C12-C16	7733315.000	60.000	6.398	9.101	128888.583	145185.730	11.225
Aromatic C16-C21	9911263.000	80.000	9.102	13.420	123890.788	144173.003	14.068
Aromatic C21-C36	20345622.000	180.000	13.421	18.876	113031.233	130834.072	13.607
Aromatic EPH	43074545.000	360.000	4.521	18.876	119651.514	137577.803	13.030