

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE092425AL

Review By	yogesh	Review On	9/24/2025 11:25:42 AM
Supervise By	mohammad	Supervise On	9/26/2025 4:25:44 AM
SubDirectory	FE092425AL	HP Acquire Method	HP Processing Method FE092425AL

STD.NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773
CCC Internal Standard/PEM	PP24771
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FE055960.D	24 Sep 2025 08:06		YP\AJ	Ok
2	I.BLK	I.BLK	FE055961.D	24 Sep 2025 08:36		YP\AJ	Ok
3	100 PPM ALIPHATIC HC	100 PPM ALIPHATIC HC	FE055962.D	24 Sep 2025 09:07		YP\AJ	Ok
4	50 PPM ALIPHATIC HC	50 PPM ALIPHATIC HC	FE055963.D	24 Sep 2025 09:37		YP\AJ	Ok
5	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE055964.D	24 Sep 2025 10:07		YP\AJ	Ok
6	10 PPM ALIPHATIC HC	10 PPM ALIPHATIC HC	FE055965.D	24 Sep 2025 10:38		YP\AJ	Ok
7	5 PPM ALIPHATIC HC	5 PPM ALIPHATIC HC	FE055966.D	24 Sep 2025 11:08		YP\AJ	Ok
8	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE055967.D	24 Sep 2025 11:39		YP\AJ	Ok
9	I.BLK	I.BLK	FE055968.D	24 Sep 2025 12:09		YP\AJ	Ok
10	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE055969.D	24 Sep 2025 12:40		YP\AJ	Ok
11	PB169809BL	PB169809BL	FE055970.D	24 Sep 2025 14:27		YP\AJ	Ok
12	PB169809BS	PB169809BLBS	FE055971.D	24 Sep 2025 14:58		YP\AJ	Ok

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE092425AL

Review By	yogesh	Review On	9/24/2025 11:25:42 AM
Supervise By	mohammad	Supervise On	9/26/2025 4:25:44 AM
SubDirectory	FE092425AL	HP Acquire Method	HP Processing Method FE092425AL
STD.NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC Internal Standard/PEM	PP24771		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774		

13	PB169809BSD	PB169809BLBSBSBSB	FE055972.D	24 Sep 2025 15:28		YPAJ	Ok
14	Q3159-01	VNJ-245	FE055973.D	24 Sep 2025 15:59		YPAJ	Ok
15	Q3159-01D	VNJ-245D	FE055974.D	24 Sep 2025 16:29		YPAJ	Ok
16	Q3159-01MS	VNJ-245MS	FE055975.D	24 Sep 2025 17:00	FE055973.D	YPAJ	Ok
17	Q3159-01MSD	VNJ-245MSD	FE055976.D	24 Sep 2025 17:31	FE055973.D!FE055975.D	YPAJ	Ok
18	Q3159-03	VNJ-255	FE055977.D	24 Sep 2025 18:01		YPAJ	Ok
19	Q3167-01	RT4784	FE055978.D	24 Sep 2025 18:32	Need 2X	YPAJ	Dilution
20	Q3168-03	ARS20-0010-Oily Soil-	FE055979.D	24 Sep 2025 19:03	Need 10X	YPAJ	Dilution
21	I.BLK	I.BLK	FE055980.D	24 Sep 2025 20:04		YPAJ	Ok
22	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE055981.D	24 Sep 2025 20:35		YPAJ	Ok

M : Manual Integration

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE092925AL

Review By	yogesh	Review On	9/29/2025 10:09:20 AM		
Supervise By	mohammad	Supervise On	10/1/2025 1:40:41 AM		
SubDirectory	FE092925AL	HP Acquire Method	HP Processing Method	FE092425AL	

STD.NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773
CCC Internal Standard/PEM	PP24771
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FE056033.D	29 Sep 2025 08:16		YP\AJ	Ok
2	I.BLK	I.BLK	FE056034.D	29 Sep 2025 08:47		YP\AJ	Ok
3	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056035.D	29 Sep 2025 09:17		YP\AJ	Ok,M
4	Q3209-01	PL-02-092525	FE056036.D	29 Sep 2025 09:47	need 10x dilution	YP\AJ	Dilution
5	Q3209-02	PL-02-092525-E2	FE056037.D	29 Sep 2025 10:18	need 5x dilution	YP\AJ	Dilution
6	Q3212-01	EO-03-092525	FE056038.D	29 Sep 2025 10:48	need 10x dilution	YP\AJ	Dilution
7	Q3212-02	EO-03-092525-E2	FE056039.D	29 Sep 2025 11:18	need 10x dilution	YP\AJ	Dilution
8	Q3209-01DL	PL-02-092525DL	FE056040.D	29 Sep 2025 11:49		YP\AJ	Ok
9	Q3209-02DL	PL-02-092525-E2DL	FE056041.D	29 Sep 2025 12:19		YP\AJ	Ok
10	Q3212-01DL	EO-03-092525DL	FE056042.D	29 Sep 2025 12:49		YP\AJ	Ok,M
11	Q3212-02DL	EO-03-092525-E2DL	FE056043.D	29 Sep 2025 13:20		YP\AJ	Ok
12	I.BLK	I.BLK	FE056044.D	29 Sep 2025 14:21		YP\AJ	Ok
13	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056045.D	29 Sep 2025 14:51		YP\AJ	Ok,M
14	Q3200-01	ETGI-367	FE056046.D	29 Sep 2025 15:22		YP\AJ	Ok
15	Q3200-02	ETGI-367	FE056047.D	29 Sep 2025 15:52		YP\AJ	Ok
16	Q3206-01	TR-06-092525	FE056048.D	29 Sep 2025 16:23		YP\AJ	Ok,M
17	Q3206-02	TR-06-092525-E2	FE056049.D	29 Sep 2025 16:53	Need 5X	YP\AJ	Dilution
18	Q3208-03	SU-ESM-01	FE056050.D	29 Sep 2025 17:23	Need 5X	YP\AJ	Dilution

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE092925AL

Review By	yogesh	Review On	9/29/2025 10:09:20 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:41 AM
SubDirectory	FE092925AL	HP Acquire Method	HP Processing Method FE092425AL
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773		
CCC Internal Standard/PEM	PP24771		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774		

19	Q3208-04	SU-ESM-02	FE056051.D	29 Sep 2025 17:54	Need 5X	YPAJ	Dilution
20	Q3177-01	Q3177-01	FE056052.D	29 Sep 2025 18:25	Possible carryover	YPAJ	Not Ok
21	Q3177-02	Q3177-02	FE056053.D	29 Sep 2025 18:55	Possible carryover	YPAJ	Not Ok
22	Q3177-03	Q3177-03	FE056054.D	29 Sep 2025 19:26	Possible carryover	YPAJ	Not Ok
23	Q3179-01	EME-TS14-01	FE056055.D	29 Sep 2025 19:56		YPAJ	Ok
24	Q3179-02MS	EME-TS14-01MS	FE056056.D	29 Sep 2025 20:27	FE056055.D	YPAJ	Ok
25	Q3179-03MSD	EME-TS14-01MSD	FE056057.D	29 Sep 2025 20:57	FE056055.D!FE056056.D	YPAJ	Ok
26	I.BLK	I.BLK	FE056058.D	29 Sep 2025 21:58		YPAJ	Ok
27	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056059.D	29 Sep 2025 22:59		YPAJ	Ok,M
28	Q3220-01	OR-03-092625	FE056060.D	29 Sep 2025 23:30		YPAJ	Ok
29	Q3220-02	OR-03-092625-E2	FE056061.D	30 Sep 2025 00:01		YPAJ	Ok
30	Q3221-01	HD-01-092625	FE056062.D	30 Sep 2025 00:31	Need 5X	YPAJ	Dilution
31	Q3221-02	HD-01-092625-E2	FE056063.D	30 Sep 2025 01:02	Need 5X	YPAJ	Dilution
32	Q3222-01	SU-03-092625	FE056064.D	30 Sep 2025 01:32	Need 5X	YPAJ	Dilution
33	Q3222-02	SU-03-092625-E2	FE056065.D	30 Sep 2025 02:02	Need 5X	YPAJ	Dilution
34	Q3231-01	WC-1	FE056066.D	30 Sep 2025 02:33	Need 2X	YPAJ	Dilution
35	Q3231-02	WC-1-EPH	FE056067.D	30 Sep 2025 03:03		YPAJ	Ok

Instrument ID: FID_E

Daily Analysis Runlog For Sequence/QC Batch ID # FE092925AL

Review By	yogesh	Review On	9/29/2025 10:09:20 AM	
Supervise By	mohammad	Supervise On	10/1/2025 1:40:41 AM	
SubDirectory	FE092925AL	HP Acquire Method	HP Processing Method	FE092425AL
STD.NAME	STD REF.#			
Tune/Reschk Initial Calibration Stds	PP24768,PP24770,PP24771,PP24772,PP24773			
CCC Internal Standard/PEM	PP24771			
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24769,PP24774			

36	Q3231-06	WC-2-EPH	FE056068.D	30 Sep 2025 03:33		YPAJ	Ok,M
37	Q3231-09	WC-3	FE056069.D	30 Sep 2025 04:04	Need 2X	YPAJ	Dilution
38	Q3231-10	WC-3-EPH	FE056070.D	30 Sep 2025 04:34	Need 2X	YPAJ	Dilution
39	Q3231-13	WC-4	FE056071.D	30 Sep 2025 05:05		YPAJ	Ok,M
40	Q3231-14	WC-4-EPH	FE056072.D	30 Sep 2025 05:35		YPAJ	Ok,M
41	PB169879BL	PB169879BL	FE056073.D	30 Sep 2025 06:05		YPAJ	Ok,M
42	PB169879BS	PB169879BS	FE056074.D	30 Sep 2025 06:36		YPAJ	Ok,M
43	PB169879TB	PB169879BSTB	FE056075.D	30 Sep 2025 07:06		YPAJ	Ok
44	I.BLK	I.BLK	FE056076.D	30 Sep 2025 07:36		YPAJ	Ok
45	20 PPM ALIPHATIC HC	20 PPM ALIPHATIC HC	FE056077.D	30 Sep 2025 08:07		YPAJ	Ok,M

M : Manual Integration

Instrument ID: FID_F

Daily Analysis Runlog For Sequence/QC Batch ID # FF092425AR

Review By	yogesh	Review On	9/24/2025 12:57:21 PM	
Supervise By	mohammad	Supervise On	9/26/2025 4:25:38 AM	
SubDirectory	FF092425AR	HP Acquire Method	HP Processing Method	FF092425AR
STD.NAME		STD REF.#		
Tune/Reschk				
Initial Calibration Stds				
CCC				
Internal Standard/PEM				
ICV/I.BLK				
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FF016429.D	24 Sep 2025 13:00		YPIAJ	Ok
2	I.BLK	I.BLK	FF016430.D	24 Sep 2025 13:30		YPIAJ	Ok
3	100 PPM AROMATIC H	100 PPM AROMATIC H	FF016431.D	24 Sep 2025 13:59		YPIAJ	Ok
4	50 PPM AROMATIC HC	50 PPM AROMATIC HC	FF016432.D	24 Sep 2025 14:29		YPIAJ	Ok
5	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FF016433.D	24 Sep 2025 14:59		YPIAJ	Ok
6	10 PPM AROMATIC HC	10 PPM AROMATIC HC	FF016434.D	24 Sep 2025 15:28		YPIAJ	Ok
7	5 PPM AROMATIC HC	5 PPM AROMATIC HC	FF016435.D	24 Sep 2025 15:58		YPIAJ	Ok
8	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FF016436.D	24 Sep 2025 16:28		YPIAJ	Ok
9	I.BLK	I.BLK	FF016437.D	24 Sep 2025 17:27		YPIAJ	Ok
10	20 PPM AROMATIC HC	20 PPM AROMATIC HC	FF016438.D	24 Sep 2025 17:57		YPIAJ	Ok

M : Manual Integration

Instrument ID: FID_F

Daily Analysis Runlog For Sequence/QC Batch ID # FF092925AR

Review By	yogesh	Review On	9/29/2025 3:49:57 PM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:54 AM
SubDirectory	FF092925AR	HP Acquire Method	HP Processing Method FF092425AR

STD.NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24500,PP24502,PP24503,PP24504,PP24505
CCC Internal Standard/PEM	PP24503
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24501,PP24506

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	MECL2	MECL2	FF016467.D	29 Sep 2025 08:51		YP\AJ	Ok
2	I.BLK	I.BLK	FF016468.D	29 Sep 2025 12:09		YP\AJ	Ok
3	20 PPM AROMATIC HC	20 PPM AROMATIC H	FF016469.D	29 Sep 2025 14:07		YP\AJ	Ok
4	PB169879BL	PB169879BL	FF016470.D	29 Sep 2025 16:35		YP\AJ	Ok
5	PB169879BS	PB169879BS	FF016471.D	29 Sep 2025 17:05		YP\AJ	Ok
6	PB169879TB	PB169879BSTB	FF016472.D	29 Sep 2025 17:35		YP\AJ	Ok
7	Q3177-01	EME-TS12-01	FF016473.D	29 Sep 2025 18:04	Surrogate Fail	YP\AJ	ReRun
8	Q3177-02	EME-TS13-01	FF016474.D	29 Sep 2025 18:34		YP\AJ	Ok
99	Q3177-03	EME-TS14-02	FF016475.D	29 Sep 2025 19:04		YP\AJ	Ok
10	Q3179-01	EME-TS14-01	FF016476.D	29 Sep 2025 19:33		YP\AJ	Ok
11	Q3179-02MS	EME-TS14-01MS	FF016477.D	29 Sep 2025 20:03	FF016476.D	YP\AJ	Ok

Instrument ID: FID_F

Daily Analysis Runlog For Sequence/QC Batch ID # FF092925AR

Review By	yogesh	Review On	9/29/2025 3:49:57 PM
Supervise By	mohammad	Supervise On	10/1/2025 1:40:54 AM
SubDirectory	FF092925AR	HP Acquire Method	HP Processing Method FF092425AR

STD.NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24500,PP24502,PP24503,PP24504,PP24505
CCC Internal Standard/PEM	PP24503
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24501,PP24506

12	Q3179-03MSD	EME-TS14-01MSD	FF016478.D	29 Sep 2025 20:33	FF016476.D\FF016477.D	YPAJ	Ok
13	I.BLK	I.BLK	FF016479.D	29 Sep 2025 21:03		YPAJ	Ok
14	20 PPM AROMATIC HC	20 PPM AROMATIC H	CFF016480.D	29 Sep 2025 22:02		YPAJ	Ok

M : Manual Integration