

WATER EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name:	<u>Alliance</u>	Client:	<u>Weston Solutions, Inc.</u>
Lab Code:	<u>ACE</u>	SDG No:	<u>Q3179</u>
Sample No :	<u>Q3179-03</u>	Datafile:	<u>FF016478.D</u>
		Client ID :	<u>EME-TS14-01MSD</u>

COMPOUND	SPIKE ADDED ug/l	SAMPLE CONCENTRATION ug/l	MS/MSD CONCENTRATION ug/l	% REC	Qual	RPD QC LIMITS (%) QC Limit Of RPD
Aromatic C10-C12	200	3.71	216	106		2.79 (40-140) 50
Aromatic C12-C16	300	7.91	359	117		16.5 (40-140) 50
Aromatic C16-C21	500	43.8	745	140		13 (40-140) 50
Aromatic C21-C36	800	0	933	116		1.71 (40-140) 50