

WATER EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name:	<u>Alliance</u>	Client:	<u>Weston Solutions, Inc.</u>
Lab Code:	<u>ACE</u>	SDG No:	<u>Q3179</u>
Sample No :	<u>PB169879BS</u>	Datafile:	<u>FF016471.D</u>
		Client ID :	<u>PB169879BS</u>

COMPOUND	SPIKE ADDED ug/l	LCS/LCSD CONCENTRATION ug/l	% REC	Qual	QC LIMITS (%)
Aromatic C10-C12	200	160	80		(40-140)
Aromatic C12-C16	300	266	89		(40-140)
Aromatic C16-C21	500	469	94		(40-140)
Aromatic C21-C36	800	747	93		(40-140)