

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075413.D **Time Analyzed:** 18:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.806	5.708	5.908	475.770	500.000	-4.8
Aroclor-1016-2	5.827	5.729	5.929	490.020	500.000	-2.0
Aroclor-1016-3	5.890	5.792	5.992	473.990	500.000	-5.2
Aroclor-1016-4	5.987	5.889	6.089	515.030	500.000	3.0
Aroclor-1016-5	6.280	6.181	6.381	518.410	500.000	3.7
Aroclor-1260-1	7.397	7.299	7.499	495.570	500.000	-0.9
Aroclor-1260-2	7.649	7.551	7.751	489.480	500.000	-2.1
Aroclor-1260-3	8.007	7.909	8.109	536.930	500.000	7.4
Aroclor-1260-4	8.235	8.137	8.337	561.600	500.000	12.3
Aroclor-1260-5	8.561	8.464	8.664	545.160	500.000	9.0
Decachlorobiphenyl	10.430	10.334	10.534	51.390	50.000	2.8
Tetrachloro-m-xylene	4.655	4.555	4.755	50.180	50.000	0.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075413.D **Time Analyzed:** 18:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.891	4.793	4.993	534.060	500.000	6.8
Aroclor-1016-2	4.950	4.851	5.051	533.950	500.000	6.8
Aroclor-1016-3	5.069	4.971	5.171	494.970	500.000	-1.0
Aroclor-1016-4	5.110	5.011	5.211	518.630	500.000	3.7
Aroclor-1016-5	5.324	5.225	5.425	511.900	500.000	2.4
Aroclor-1260-1	6.352	6.255	6.455	484.640	500.000	-3.1
Aroclor-1260-2	6.539	6.441	6.641	484.970	500.000	-3.0
Aroclor-1260-3	6.692	6.595	6.795	482.320	500.000	-3.5
Aroclor-1260-4	7.160	7.065	7.265	454.900	500.000	-9.0
Aroclor-1260-5	7.400	7.304	7.504	489.010	500.000	-2.2
Decachlorobiphenyl	8.798	8.703	8.903	58.950	50.000	17.9
Tetrachloro-m-xylene	3.796	3.696	3.896	53.900	50.000	7.8