

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075428.D **Time Analyzed:** 23:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.807	5.708	5.908	488.120	500.000	-2.4
Aroclor-1016-2	5.828	5.729	5.929	496.430	500.000	-0.7
Aroclor-1016-3	5.891	5.792	5.992	486.040	500.000	-2.8
Aroclor-1016-4	5.988	5.889	6.089	519.590	500.000	3.9
Aroclor-1016-5	6.281	6.181	6.381	527.540	500.000	5.5
Aroclor-1260-1	7.398	7.299	7.499	507.550	500.000	1.5
Aroclor-1260-2	7.650	7.551	7.751	502.340	500.000	0.5
Aroclor-1260-3	8.008	7.909	8.109	547.800	500.000	9.6
Aroclor-1260-4	8.237	8.137	8.337	575.600	500.000	15.1
Aroclor-1260-5	8.562	8.464	8.664	563.440	500.000	12.7
Decachlorobiphenyl	10.431	10.334	10.534	53.400	50.000	6.8
Tetrachloro-m-xylene	4.655	4.555	4.755	50.830	50.000	1.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075428.D **Time Analyzed:** 23:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.892	4.793	4.993	537.660	500.000	7.5
Aroclor-1016-2	4.950	4.851	5.051	549.500	500.000	9.9
Aroclor-1016-3	5.070	4.971	5.171	497.520	500.000	-0.5
Aroclor-1016-4	5.111	5.011	5.211	525.220	500.000	5.0
Aroclor-1016-5	5.324	5.225	5.425	505.020	500.000	1.0
Aroclor-1260-1	6.352	6.255	6.455	506.680	500.000	1.3
Aroclor-1260-2	6.539	6.441	6.641	496.240	500.000	-0.8
Aroclor-1260-3	6.692	6.595	6.795	511.100	500.000	2.2
Aroclor-1260-4	7.161	7.065	7.265	468.620	500.000	-6.3
Aroclor-1260-5	7.401	7.304	7.504	496.780	500.000	-0.6
Decachlorobiphenyl	8.796	8.703	8.903	61.070	50.000	22.1
Tetrachloro-m-xylene	3.797	3.696	3.896	54.520	50.000	9.0