

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075449.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.808	5.708	5.908	473.340	500.000	-5.3
Aroclor-1016-2	5.829	5.729	5.929	477.380	500.000	-4.5
Aroclor-1016-3	5.892	5.792	5.992	463.180	500.000	-7.4
Aroclor-1016-4	5.989	5.889	6.089	498.870	500.000	-0.2
Aroclor-1016-5	6.281	6.181	6.381	501.220	500.000	0.2
Aroclor-1260-1	7.399	7.299	7.499	487.040	500.000	-2.6
Aroclor-1260-2	7.651	7.551	7.751	475.980	500.000	-4.8
Aroclor-1260-3	8.010	7.909	8.109	518.370	500.000	3.7
Aroclor-1260-4	8.238	8.137	8.337	545.630	500.000	9.1
Aroclor-1260-5	8.564	8.464	8.664	541.150	500.000	8.2
Decachlorobiphenyl	10.434	10.334	10.534	52.310	50.000	4.6
Tetrachloro-m-xylene	4.656	4.555	4.755	50.130	50.000	0.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3179
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075449.D **Time Analyzed:** 16:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.891	4.793	4.993	542.910	500.000	8.6
Aroclor-1016-2	4.950	4.851	5.051	569.090	500.000	13.8
Aroclor-1016-3	5.070	4.971	5.171	505.350	500.000	1.1
Aroclor-1016-4	5.111	5.011	5.211	533.050	500.000	6.6
Aroclor-1016-5	5.324	5.225	5.425	531.740	500.000	6.3
Aroclor-1260-1	6.353	6.255	6.455	423.560	500.000	-15.3
Aroclor-1260-2	6.539	6.441	6.641	428.070	500.000	-14.4
Aroclor-1260-3	6.692	6.595	6.795	439.130	500.000	-12.2
Aroclor-1260-4	7.161	7.065	7.265	415.220	500.000	-17.0
Aroclor-1260-5	7.401	7.304	7.504	432.220	500.000	-13.6
Decachlorobiphenyl	8.799	8.703	8.903	56.660	50.000	13.3
Tetrachloro-m-xylene	3.796	3.696	3.896	54.770	50.000	9.5