

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3179</u>	Analytical Method:	<u>8082A</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PP075423.D</u>

	Parameter	Sample Spike	Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3179-03MSD (Column 1)	Client Sample ID:	EME-TS14-01MSD									
	AR1016	5	0	6.70	ug/L	134		3		66	146	20
	AR1260	5	0	6.90	ug/L	138		3		49	147	20
Lab Sample ID:	Q3179-03MSD (Column 2)	Client Sample ID:	EME-TS14-01MSD									
	AR1016	5	0	6.80	ug/L	136		4		66	146	20
	AR1260	5	0	6.50	ug/L	130		6		49	147	20