

Surrogate Summary

SDG No.: Q3179

Client: Weston Solutions, Inc.

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PD090327.D	PIBLK-PD090327.D	Decachlorobiphen	1	20	17.6	88		57	171
		Tetrachloro-m-xyl	1	20	16.5	83		61	148
		Decachlorobiphen	2	20	17.2	86		57	171
		Tetrachloro-m-xyl	2	20	16.9	85		61	148
I.BLK-PD090416.D	PIBLK-PD090416.D	Decachlorobiphen	1	20	18.2	91		57	171
		Tetrachloro-m-xyl	1	20	19.3	97		61	148
		Decachlorobiphen	2	20	17.3	87		57	171
		Tetrachloro-m-xyl	2	20	20.4	102		61	148
PB169882BL	PB169882BL	Decachlorobiphen	1	20	27.5	137		57	171
		Tetrachloro-m-xyl	1	20	26.2	131		61	148
		Decachlorobiphen	2	20	26.0	130		57	171
		Tetrachloro-m-xyl	2	20	26.9	134		61	148
PB169882BS	PB169882BS	Decachlorobiphen	1	20	23.2	116		57	171
		Tetrachloro-m-xyl	1	20	23.9	120		61	148
		Decachlorobiphen	2	20	22.4	112		57	171
		Tetrachloro-m-xyl	2	20	23.8	119		61	148
PB169882TB	PB169882TB	Decachlorobiphen	1	20	14.5	72		57	171
		Tetrachloro-m-xyl	1	20	15.5	78		61	148
		Decachlorobiphen	2	20	14.1	70		57	171
		Tetrachloro-m-xyl	2	20	15.2	76		61	148
Q3179-01	EME-TS14-01	Decachlorobiphen	1	20	14.9	74		57	171
		Tetrachloro-m-xyl	1	20	22.0	110		61	148
		Decachlorobiphen	2	20	13.1	65		57	171
		Tetrachloro-m-xyl	2	20	20.5	102		61	148
Q3179-02MS	EME-TS14-01MS	Decachlorobiphen	1	20	15.5	78		57	171
		Tetrachloro-m-xyl	1	20	17.7	88		61	148
		Decachlorobiphen	2	20	13.9	70		57	171
		Tetrachloro-m-xyl	2	20	16.9	84		61	148
Q3179-03MSD	EME-TS14-01MSD	Decachlorobiphen	1	20	14.8	74		57	171
		Tetrachloro-m-xyl	1	20	17.7	88		61	148
		Decachlorobiphen	2	20	13.6	68		57	171
		Tetrachloro-m-xyl	2	20	16.6	83		61	148
I.BLK-PD090431.D	PIBLK-PD090431.D	Decachlorobiphen	1	20	19.8	99		57	171
		Tetrachloro-m-xyl	1	20	20.1	100		61	148
		Decachlorobiphen	2	20	19.1	96		57	171
		Tetrachloro-m-xyl	2	20	19.7	99		61	148