

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3179</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PD090429.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
			Result								High	RPD
Lab Sample ID:	Q3179-02MS (Column 1)	Client Sample ID:		EME-TS14-01MS								
	alpha-BHC	0.05	0	0.042	ug/L	83				46	160	
	beta-BHC	0.05	0	0.040	ug/L	79				15	175	
	delta-BHC	0.05	0	0.046	ug/L	92				10	175	
	gamma-BHC (Lindane)	0.05	0	0.045	ug/L	90				60	152	
	Heptachlor	0.05	0	0.038	ug/L	75				56	147	
	Aldrin	0.05	0	0.036	ug/L	72				45	147	
	Heptachlor epoxide	0.05	0	0.041	ug/L	82				77	143	
	Endosulfan I	0.05	0	0.042	ug/L	84				34	157	
	Dieldrin	0.05	0	0.043	ug/L	85				46	155	
	4,4'-DDE	0.05	0	0.041	ug/L	82				36	162	
	Endrin	0.05	0	0.044	ug/L	87				76	144	
	Endosulfan II	0.05	0	0.043	ug/L	87				21	168	
	4,4'-DDD	0.05	0	0.042	ug/L	84				15	175	
	Endosulfan sulfate	0.05	0	0.043	ug/L	85				14	183	
	4,4'-DDT	0.05	0	0.044	ug/L	87				15	175	
	Methoxychlor	0.05	0	0.045	ug/L	90				70	142	
	Endrin ketone	0.05	0	0.044	ug/L	87				25	172	
	Endrin aldehyde	0.05	0	0.044	ug/L	88				28	175	
	alpha-Chlordane	0.05	0	0.040	ug/L	80				34	160	
	gamma-Chlordane	0.05	0	0.039	ug/L	78				31	163	
Lab Sample ID:	Q3179-02MS (Column 2)	Client Sample ID:		EME-TS14-01MS								
	alpha-BHC	0.05	0	0.040	ug/L	79				46	160	
	beta-BHC	0.05	0	0.042	ug/L	83				15	175	
	delta-BHC	0.05	0	0.039	ug/L	78				10	175	
	gamma-BHC (Lindane)	0.05	0	0.039	ug/L	78				60	152	
	Heptachlor	0.05	0	0.037	ug/L	74				56	147	
	Aldrin	0.05	0	0.036	ug/L	71				45	147	
	Heptachlor epoxide	0.05	0	0.042	ug/L	84				77	143	
	Endosulfan I	0.05	0	0.040	ug/L	80				34	157	
	Dieldrin	0.05	0	0.040	ug/L	80				46	155	
	4,4'-DDE	0.05	0	0.038	ug/L	77				36	162	
	Endrin	0.05	0	0.040	ug/L	80				76	144	
	Endosulfan II	0.05	0	0.040	ug/L	80				21	168	
	4,4'-DDD	0.05	0	0.039	ug/L	78				15	175	
	Endosulfan sulfate	0.05	0	0.040	ug/L	80				14	183	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q3179</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Weston Solutions, Inc.</u>	DataFile :	<u>PD090429.D</u>

Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits	
		Result	Result							High	RPD
4,4'-DDT	0.05	0	0.040	ug/L	80				15	175	
Methoxychlor	0.05	0	0.040	ug/L	80				70	142	
Endrin ketone	0.05	0	0.038	ug/L	77				25	172	
Endrin aldehyde	0.05	0	0.042	ug/L	83				28	175	
alpha-Chlordane	0.05	0	0.039	ug/L	78				34	160	
gamma-Chlordane	0.05	0	0.039	ug/L	78				31	163	