

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3179

Analytical Method: 8081B

Client: Weston Solutions, Inc.

Datafile : PD090423.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB169882BS (Column 1)	alpha-BHC	0.05	0.049	ug/L	98				85	130	
	beta-BHC	0.05	0.046	ug/L	92				83	126	
	delta-BHC	0.05	0.049	ug/L	98				69	141	
	gamma-BHC (Lindane)	0.05	0.048	ug/L	96				82	129	
	Heptachlor	0.05	0.050	ug/L	99				79	127	
	Aldrin	0.05	0.049	ug/L	97				79	126	
	Heptachlor epoxide	0.05	0.048	ug/L	96				81	124	
	Endosulfan I	0.05	0.048	ug/L	95				85	122	
	Dieldrin	0.05	0.049	ug/L	98				83	125	
	4,4'-DDE	0.05	0.049	ug/L	97				80	127	
	Endrin	0.05	0.051	ug/L	101				81	128	
	Endosulfan II	0.05	0.050	ug/L	101				82	123	
	4,4'-DDD	0.05	0.051	ug/L	102				77	131	
	Endosulfan sulfate	0.05	0.048	ug/L	96				76	129	
	4,4'-DDT	0.05	0.049	ug/L	97				80	133	
	Methoxychlor	0.05	0.053	ug/L	107				78	108	
	Endrin ketone	0.05	0.048	ug/L	96				80	131	
	Endrin aldehyde	0.05	0.051	ug/L	103				82	127	
	alpha-Chlordane	0.05	0.045	ug/L	89				82	125	
	gamma-Chlordane	0.05	0.046	ug/L	92				82	125	
	alpha-BHC	0.05	0.047	ug/L	93				85	130	
PB169882BS (Column 2)	beta-BHC	0.05	0.046	ug/L	92				83	126	
	delta-BHC	0.05	0.046	ug/L	93				69	141	
	gamma-BHC (Lindane)	0.05	0.047	ug/L	93				82	129	
	Heptachlor	0.05	0.047	ug/L	94				79	127	
	Aldrin	0.05	0.047	ug/L	93				79	126	
	Heptachlor epoxide	0.05	0.057	ug/L	113				81	124	
	Endosulfan I	0.05	0.044	ug/L	88				85	122	
	Dieldrin	0.05	0.046	ug/L	91				83	125	
	4,4'-DDE	0.05	0.047	ug/L	93				80	127	
	Endrin	0.05	0.050	ug/L	99				81	128	
	Endosulfan II	0.05	0.046	ug/L	92				82	123	
	4,4'-DDD	0.05	0.046	ug/L	92				77	131	
	Endosulfan sulfate	0.05	0.046	ug/L	91				76	129	
	4,4'-DDT	0.05	0.046	ug/L	93				80	133	
	Methoxychlor	0.05	0.046	ug/L	93				78	108	
	Endrin ketone	0.05	0.051	ug/L	103				80	131	
	Endrin aldehyde	0.05	0.048	ug/L	97				82	127	
	alpha-Chlordane	0.05	0.046	ug/L	92				82	125	
	gamma-Chlordane	0.05	0.046	ug/L	92				82	125	