

Surrogate Summary

SW-846

SDG No.: Q3179

Client: Weston Solutions, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB169880BL	PB169880BL	2-Fluorophenol	150	103	68		23	138
		Phenol-d6	150	105	70		10	134
		Nitrobenzene-d5	100	74.1	74		67	132
		2-Fluorobiphenyl	100	73.6	74		52	132
		2,4,6-Tribromophenol	150	109	73		44	137
		Terphenyl-d14	100	91.7	92		42	152
PB169880BS	PB169880BS	2-Fluorophenol	150	108	72		23	138
		Phenol-d6	150	112	74		10	134
		Nitrobenzene-d5	100	82.5	82		67	132
		2-Fluorobiphenyl	100	78.5	78		52	132
		2,4,6-Tribromophenol	150	109	72		44	137
		Terphenyl-d14	100	96.8	97		42	152
PB169880TB	PB169880TB	2-Fluorophenol	150	109	73		23	138
		Phenol-d6	150	111	74		10	134
		Nitrobenzene-d5	100	82.1	82		67	132
		2-Fluorobiphenyl	100	79.2	79		52	132
		2,4,6-Tribromophenol	150	114	76		44	137
		Terphenyl-d14	100	94.6	95		42	152
Q3179-01	EME-TS14-01	2-Fluorophenol	150	58.3	39		23	138
		Phenol-d6	150	38.4	26		10	134
		Nitrobenzene-d5	100	92.8	93		67	132
		2-Fluorobiphenyl	100	85.7	86		52	132
		2,4,6-Tribromophenol	150	139	92		44	137
		Terphenyl-d14	100	99.3	99		42	152
Q3179-02MS	EME-TS14-01MS	2-Fluorophenol	150	60.4	40		23	138
		Phenol-d6	150	41.0	27		10	134
		Nitrobenzene-d5	100	92.4	92		67	132
		2-Fluorobiphenyl	100	85.4	85		52	132
		2,4,6-Tribromophenol	150	129	86		44	137
		Terphenyl-d14	100	109	109		42	152
Q3179-03MSD	EME-TS14-01MSD	2-Fluorophenol	150	61.6	41		23	138
		Phenol-d6	150	41.8	28		10	134
		Nitrobenzene-d5	100	94.9	95		67	132
		2-Fluorobiphenyl	100	86.9	87		52	132
		2,4,6-Tribromophenol	150	130	87		44	137
		Terphenyl-d14	100	112	112		42	152