

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name:	<u>Alliance</u>	Contract:	<u>ROYF02</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3179</u>
Lab File ID:	<u>VN087948.D</u>	Date Analyzed:	<u>09/25/2025</u>
Instrument ID:	<u>MSVOA_N</u>	Time Analyzed:	<u>12:04</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Heated Purge:	(Y/N) <u>N</u>

	IS4 AREA #	RT #				
12 HOUR STD	240530	13.77				
UPPER LIMIT	481060	14.27				
LOWER LIMIT	120265	13.27				
EPA SAMPLE NO.						
EME-TS14-01	166506	13.77				
EME-TS14-01MS	221292	13.77				
EME-TS14-01MSD	236701	13.77				
VN0925WBL01	171513	13.77				
VN0925WBS01	244468	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.