

Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB137323

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3178
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3191
buffer solution pH 7 yellow	W3217
Buffer Solution, PH2 (500ml)	W3161
pH 12.00 Buffer	W3200

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.05].

True Value of CCV2 = 12.00 Control Limits[+/- 0.05].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	09/26/2025	08:00
2	CAL2	1	Water	NA	NA	20.3	7.01	09/26/2025	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	09/26/2025	08:15
4	ICV	1	Water	NA	NA	20.3	7.00	09/26/2025	08:22
5	CCV1	1	Water	NA	NA	20.3	2.01	09/26/2025	08:30
6	Q3181-02	1	Solid	20.02	20	24.5	11.24	09/26/2025	08:40
7	Q3181-02DUP	1	Solid	20.03	20	24.6	11.26	09/26/2025	08:45
8	Q3181-06	1	Solid	20.02	20	24.1	11.39	09/26/2025	08:55
9	Q3181-10	1	Solid	20.04	20	24.7	11.74	09/26/2025	09:00
10	CCV2	1	Water	NA	NA	20.2	12.02	09/26/2025	09:05

WORKLIST(Hardcopy Internal Chain)

VP 1371323

WorkList Name : ph s q3181 WorkList ID : 192104 Department : Wet-Chemistry Date : 09-26-2025 07:46:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3181-02	WC-A3-01-C	Solid	pH	Cool 4 deg C	ENTA05	J33	09/24/2025	9045D
Q3181-06	WC-A3-02-C	Solid	pH	Cool 4 deg C	ENTA05	J33	09/24/2025	9045D
Q3181-10	WC-A3-03-C	Solid	pH	Cool 4 deg C	ENTA05	J33	09/24/2025	9045D

Date/Time 09/26/25 07:55
Raw Sample Received by: SP WC/
Raw Sample Relinquished by: AR S

Date/Time 09/26/25 13:00
Raw Sample Received by: CR S
Raw Sample Relinquished by: SP WC/