



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3181

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB137322

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.00	7	100	90-110	09/26/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	09/26/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	09/26/2025
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	09/26/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3181

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB137323

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	09/26/2025
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	09/26/2025
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	09/26/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3181

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB137332

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.095	0.099	96	85-115	09/26/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/26/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.23	0.25	92	90-110	09/26/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/26/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/26/2025
Sample ID: CCV5 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	09/26/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3181

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB137347

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 ASTM Ammonia	mg/L	0.96	1	96	90-110	09/29/2025
Sample ID: CCV1 ASTM Ammonia	mg/L	1	1	100	90-110	09/29/2025
Sample ID: CCV2 ASTM Ammonia	mg/L	1	1	100	90-110	09/29/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3181

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB137350

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV ASTM COD	mg/L	50.173	50	100	95-105	09/08/2025
Sample ID: CCV1 ASTM COD	mg/L	49.161	50	98	95-105	09/29/2025
Sample ID: CCV2 ASTM COD	mg/L	49.161	50	98	95-105	09/29/2025

Initial and Continuing Calibration Verification

Client: ENTACT

SDG No.: Q3181

Project: 540 Degraw St, Brooklyn, NY - E9309

RunNo.: LB137350

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------