



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT

SDG No.: Q3181

Contract: ENTA05

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	87	70 - 130	CV	09/29/2025	10:56	LB137343
CRI01	Arsenic	21.9	20.0	110	65 - 135	P	09/29/2025	13:08	LB137351
	Barium	97.6	100	98	65 - 135	P	09/29/2025	13:08	LB137351
	Cadmium	6.00	6.0	100	65 - 135	P	09/29/2025	13:08	LB137351
	Chromium	9.91	10.0	99	65 - 135	P	09/29/2025	13:08	LB137351
	Copper	21.0	20.0	105	65 - 135	P	09/29/2025	13:08	LB137351
	Lead	13.2	12.0	110	65 - 135	P	09/29/2025	13:08	LB137351
	Nickel	41.0	40.0	102	65 - 135	P	09/29/2025	13:08	LB137351
	Selenium	19.7	20.0	98	65 - 135	P	09/29/2025	13:08	LB137351
	Silver	9.87	10.0	99	65 - 135	P	09/29/2025	13:08	LB137351
	Zinc	41.0	40.0	103	65 - 135	P	09/29/2025	13:08	LB137351