

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q3181

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169840TB		WATER		Batch Number:	PB169867		Prep Date:	09/26/2025	
	Mercury	0.76	<2	U	2.00	CV	09/29/2025	13:00	LB137343
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169867BL		WATER		Batch Number:	PB169867		Prep Date:	09/26/2025	
	Mercury	0.076	<0.2	U	0.20	CV	09/29/2025	11:03	LB137343

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q3181

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169840TB		WATER		Batch Number:	PB169865		Prep Date:	09/26/2025	
	Arsenic	25.6	<50	U	100	P	09/29/2025	16:22	LB137351
	Barium	221	<250	J	500	P	09/29/2025	16:22	LB137351
	Cadmium	2.50	<15	U	30.0	P	09/29/2025	16:22	LB137351
	Chromium	10.6	<25	U	50.0	P	09/29/2025	16:22	LB137351
	Copper	23.0	<50	U	100	P	09/29/2025	16:22	LB137351
	Lead	11.5	<30	U	60.0	P	09/29/2025	16:22	LB137351
	Nickel	15.3	<100	U	200	P	09/29/2025	16:22	LB137351
	Selenium	48.2	<50	U	100	P	09/29/2025	16:22	LB137351
	Silver	8.10	<25	U	50.0	P	09/29/2025	16:22	LB137351
	Zinc	97.4	<100	J	200	P	09/29/2025	16:22	LB137351
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB169865BL		WATER		Batch Number:	PB169865		Prep Date:	09/26/2025	
	Arsenic	25.6	<50	U	100	P	09/29/2025	15:21	LB137351
	Barium	72.8	<250	U	500	P	09/29/2025	15:21	LB137351
	Cadmium	2.50	<15	U	30.0	P	09/29/2025	15:21	LB137351
	Chromium	10.6	<25	U	50.0	P	09/29/2025	15:21	LB137351
	Copper	23.0	<50	U	100	P	09/29/2025	15:21	LB137351
	Lead	11.5	<30	U	60.0	P	09/29/2025	15:21	LB137351
	Nickel	15.3	<100	U	200	P	09/29/2025	15:21	LB137351
	Selenium	48.2	<50	U	100	P	09/29/2025	15:21	LB137351
	Silver	8.10	<25	U	50.0	P	09/29/2025	15:21	LB137351
	Zinc	83.3	<100	U	200	P	09/29/2025	15:21	LB137351