

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>ENTACT</u>	SDG No.:	<u>Q3181</u>
Contract:	<u>ENTA05</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.59			-20	20	09/29/2025	13:13	LB137351
	Barium	2.06	6.0	34	-94	106	09/29/2025	13:13	LB137351
	Cadmium	1.18	1.0	118	-5	7	09/29/2025	13:13	LB137351
	Chromium	48.3	52.0	93	42	62	09/29/2025	13:13	LB137351
	Copper	21.1	2.0	1055	-18	22	09/29/2025	13:13	LB137351
	Lead	0.53			-12	12	09/29/2025	13:13	LB137351
	Nickel	3.83	2.0	192	-38	42	09/29/2025	13:13	LB137351
	Selenium	3.75			-20	20	09/29/2025	13:13	LB137351
	Silver	0.65			-10	10	09/29/2025	13:13	LB137351
	Zinc	1.67			-40	40	09/29/2025	13:13	LB137351
ICSAB01	Arsenic	112	104	108	88.4	120	09/29/2025	13:17	LB137351
	Barium	502	537	94	437	637	09/29/2025	13:17	LB137351
	Cadmium	993	972	102	826	1120	09/29/2025	13:17	LB137351
	Chromium	549	542	101	460	624	09/29/2025	13:17	LB137351
	Copper	507	511	99	434	588	09/29/2025	13:17	LB137351
	Lead	49.8	49.0	102	37	61	09/29/2025	13:17	LB137351
	Nickel	991	954	104	810	1100	09/29/2025	13:17	LB137351
	Selenium	57.4	46.0	125	26	66	09/29/2025	13:17	LB137351
	Silver	207	201	103	170	232	09/29/2025	13:17	LB137351
	Zinc	1020	952	107	809	1095	09/29/2025	13:17	LB137351