

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: ENTACT

SDG No.: Q3181

Contract: ENTA05

Lab Code: ACE

Matrix: Water

Level: LOW

Client ID: SU-ESM-COMP-02A

Sample ID: Q3208-12

Spiked ID: Q3208-12A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Barium	ug/L	75 - 125	1380		720		1000	67	N	P
Zinc	ug/L	75 - 125	1560		739		1000	82		P