

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q3181
Contract: ENTA05 **Lab Code:** ACE
Matrix: Water **Sample ID:** Q3208-12 **Client ID:** SU-ESM-COMP-02DUP
Percent Solids for Sample: NA **Duplicate ID** Q3208-12DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	720		847		16		P
Cadmium	ug/L	20	30.0	U	3.55	J	200.0		P
Chromium	ug/L	20	14.5	J	18.1	J	22		P
Copper	ug/L	20	71.4	J	86.6	J	19		P
Lead	ug/L	20	289		332		14		P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	15.3	J	17.9	J	16		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	739		870		16		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>ENTACT</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3181</u>
Contract:	<u>ENTA05</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3208-12MS</u>	Client ID:	<u>SU-ESM-COMP-02MSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3208-12MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4060		3600		12		P
Barium	ug/L	20	2140		1550		32	*	P
Cadmium	ug/L	20	1010		884		13		P
Chromium	ug/L	20	2060		1810		13		P
Copper	ug/L	20	1580		1370		14		P
Lead	ug/L	20	5260		4530		15		P
Mercury	ug/L	20	42.1		43.6		4		CV
Nickel	ug/L	20	2550		2230		13		P
Selenium	ug/L	20	10200		8990		13		P
Silver	ug/L	20	364		317		14		P
Zinc	ug/L	20	2210		1610		31	*	P