

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ENTA05</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3181</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):	<u>09/23/2025</u> <u>09/23/2025</u>
		Calibration Times:	<u>08:45</u> <u>17:14</u>

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP075247.D</u>	RT 750 = <u>PP075248.D</u>
RT 500 = <u>PP075249.D</u>	RT 250 = <u>PP075250.D</u>	RT 050 = <u>PP075251.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.81	5.81	5.81	5.81	5.82	5.81	5.71	5.91
Aroclor-1016-2	(2)	5.83	5.83	5.83	5.83	5.84	5.83	5.73	5.93
Aroclor-1016-3	(3)	5.89	5.90	5.89	5.89	5.90	5.89	5.79	5.99
Aroclor-1016-4	(4)	5.99	5.99	5.99	5.99	6.00	5.99	5.89	6.09
Aroclor-1016-5	(5)	6.28	6.29	6.28	6.28	6.29	6.28	6.18	6.38
Aroclor-1260-1	(1)	7.40	7.40	7.40	7.40	7.41	7.40	7.30	7.50
Aroclor-1260-2	(2)	7.65	7.66	7.65	7.65	7.66	7.65	7.55	7.75
Aroclor-1260-3	(3)	8.01	8.01	8.01	8.01	8.02	8.01	7.91	8.11
Aroclor-1260-4	(4)	8.24	8.24	8.24	8.24	8.25	8.24	8.14	8.34
Aroclor-1260-5	(5)	8.56	8.57	8.56	8.56	8.57	8.57	8.47	8.67
Decachlorobiphenyl		10.43	10.44	10.43	10.43	10.45	10.44	10.34	10.54
Tetrachloro-m-xylene		4.65	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1242-1	(1)	5.81	5.81	5.81	5.81	5.81	5.81	5.71	5.91
Aroclor-1242-2	(2)	5.83	5.83	5.83	5.83	5.83	5.83	5.73	5.93
Aroclor-1242-3	(3)	5.90	5.89	5.90	5.90	5.90	5.90	5.80	6.00
Aroclor-1242-4	(4)	5.99	5.99	5.99	5.99	5.99	5.99	5.89	6.09
Aroclor-1242-5	(5)	6.72	6.72	6.72	6.72	6.72	6.72	6.62	6.82
Decachlorobiphenyl		10.44	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1248-1	(1)	5.81	5.81	5.81	5.81	5.81	5.81	5.71	5.91
Aroclor-1248-2	(2)	6.08	6.08	6.08	6.08	6.08	6.08	5.98	6.18
Aroclor-1248-3	(3)	6.28	6.28	6.29	6.28	6.28	6.28	6.18	6.38
Aroclor-1248-4	(4)	6.68	6.68	6.68	6.68	6.68	6.68	6.58	6.78
Aroclor-1248-5	(5)	6.72	6.72	6.72	6.72	6.72	6.72	6.62	6.82
Decachlorobiphenyl		10.43	10.44	10.44	10.44	10.44	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1254-1	(1)	6.66	6.66	6.66	6.66	6.66	6.66	6.56	6.76
Aroclor-1254-2	(2)	6.88	6.87	6.88	6.88	6.87	6.87	6.77	6.97
Aroclor-1254-3	(3)	7.24	7.24	7.24	7.24	7.23	7.24	7.14	7.34
Aroclor-1254-4	(4)	7.52	7.52	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1254-5	(5)	7.94	7.93	7.94	7.94	7.93	7.93	7.83	8.03
Decachlorobiphenyl		10.44	10.44	10.44	10.44	10.43	10.44	10.34	10.54
Tetrachloro-m-xylene		4.66	4.66	4.66	4.66	4.66	4.66	4.56	4.76
Aroclor-1268-1	(1)	8.89	8.89	8.89	8.89	8.89	8.89	8.79	8.99
Aroclor-1268-2	(2)	8.99	8.99	8.99	8.98	8.98	8.99	8.89	9.09
Aroclor-1268-3	(3)	9.23	9.23	9.23	9.23	9.22	9.23	9.13	9.33
Aroclor-1268-4	(4)	9.66	9.66	9.66	9.65	9.65	9.65	9.55	9.75
Aroclor-1268-5	(5)	10.09	10.09	10.09	10.09	10.08	10.09	9.99	10.19

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ENTA05</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q3181</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):	<u>09/23/2025</u> <u>09/23/2025</u>
		Calibration Times:	<u>08:45</u> <u>17:14</u>

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP075247.D</u>	RT 750 = <u>PP075248.D</u>
RT 500 = <u>PP075249.D</u>	RT 250 = <u>PP075250.D</u>	RT 050 = <u>PP075251.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.90	4.89	4.89	4.89	4.90	4.90	4.80	5.00
Aroclor-1016-2	(2)	4.95	4.95	4.95	4.95	4.95	4.95	4.85	5.05
Aroclor-1016-3	(3)	5.07	5.07	5.07	5.07	5.07	5.07	4.97	5.17
Aroclor-1016-4	(4)	5.11	5.11	5.11	5.11	5.12	5.11	5.01	5.21
Aroclor-1016-5	(5)	5.33	5.33	5.33	5.33	5.33	5.33	5.23	5.43
Aroclor-1260-1	(1)	6.36	6.36	6.36	6.35	6.36	6.36	6.26	6.46
Aroclor-1260-2	(2)	6.54	6.54	6.54	6.54	6.55	6.54	6.44	6.64
Aroclor-1260-3	(3)	6.70	6.70	6.70	6.70	6.70	6.70	6.60	6.80
Aroclor-1260-4	(4)	7.17	7.17	7.17	7.16	7.17	7.17	7.07	7.27
Aroclor-1260-5	(5)	7.41	7.40	7.40	7.40	7.41	7.41	7.31	7.51
Decachlorobiphenyl		8.81	8.80	8.80	8.80	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1242-1	(1)	4.90	4.90	4.90	4.89	4.89	4.90	4.80	5.00
Aroclor-1242-2	(2)	4.95	4.96	4.95	4.95	4.95	4.95	4.85	5.05
Aroclor-1242-3	(3)	5.07	5.07	5.07	5.07	5.08	5.07	4.97	5.17
Aroclor-1242-4	(4)	5.16	5.16	5.16	5.15	5.16	5.16	5.06	5.26
Aroclor-1242-5	(5)	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Decachlorobiphenyl		8.81	8.81	8.81	8.81	8.80	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1248-1	(1)	4.89	4.90	4.89	4.90	4.90	4.90	4.80	5.00
Aroclor-1248-2	(2)	5.11	5.12	5.11	5.12	5.11	5.11	5.01	5.21
Aroclor-1248-3	(3)	5.15	5.16	5.15	5.16	5.15	5.16	5.06	5.26
Aroclor-1248-4	(4)	5.33	5.33	5.33	5.33	5.33	5.33	5.23	5.43
Aroclor-1248-5	(5)	5.72	5.72	5.71	5.72	5.73	5.72	5.62	5.82
Decachlorobiphenyl		8.80	8.81	8.80	8.81	8.80	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1254-1	(1)	5.68	5.68	5.68	5.68	5.68	5.68	5.58	5.78
Aroclor-1254-2	(2)	5.83	5.83	5.83	5.83	5.83	5.83	5.73	5.93
Aroclor-1254-3	(3)	6.23	6.23	6.23	6.23	6.23	6.23	6.13	6.33
Aroclor-1254-4	(4)	6.45	6.45	6.46	6.45	6.45	6.45	6.35	6.55
Aroclor-1254-5	(5)	6.87	6.87	6.87	6.87	6.87	6.87	6.77	6.97
Decachlorobiphenyl		8.81	8.81	8.81	8.81	8.81	8.81	8.71	8.91
Tetrachloro-m-xylene		3.80	3.80	3.80	3.80	3.80	3.80	3.70	3.90
Aroclor-1268-1	(1)	7.69	7.69	7.69	7.69	7.69	7.69	7.59	7.79
Aroclor-1268-2	(2)	7.75	7.75	7.75	7.75	7.75	7.75	7.65	7.85
Aroclor-1268-3	(3)	7.96	7.96	7.96	7.96	7.96	7.96	7.86	8.06
Aroclor-1268-4	(4)	8.25	8.25	8.25	8.25	8.25	8.25	8.15	8.35
Aroclor-1268-5	(5)	8.55	8.55	8.55	8.54	8.54	8.54	8.44	8.64

[illegible]