

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114080.D **Time Analyzed:** 09:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.738	4.643	4.843	496.860	500.000	-0.6
Aroclor-1016-2	4.756	4.661	4.861	508.350	500.000	1.7
Aroclor-1016-3	4.813	4.718	4.918	488.010	500.000	-2.4
Aroclor-1016-4	4.932	4.837	5.037	475.730	500.000	-4.9
Aroclor-1016-5	5.188	5.093	5.293	521.540	500.000	4.3
Aroclor-1260-1	6.223	6.129	6.329	528.750	500.000	5.8
Aroclor-1260-2	6.413	6.319	6.519	505.770	500.000	1.2
Aroclor-1260-3	6.777	6.683	6.883	511.560	500.000	2.3
Aroclor-1260-4	7.036	6.941	7.141	570.060	500.000	14.0
Aroclor-1260-5	7.279	7.185	7.385	592.980	500.000	18.6
Decachlorobiphenyl	8.662	8.570	8.770	54.030	50.000	8.1
Tetrachloro-m-xylene	3.657	3.560	3.760	53.580	50.000	7.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/15/2025 09/15/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO114080.D **Time Analyzed:** 09:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.717	4.623	4.823	456.000	500.000	-8.8
Aroclor-1016-2	4.735	4.641	4.841	455.140	500.000	-9.0
Aroclor-1016-3	4.910	4.816	5.016	454.740	500.000	-9.1
Aroclor-1016-4	4.952	4.858	5.058	463.410	500.000	-7.3
Aroclor-1016-5	5.164	5.070	5.270	476.700	500.000	-4.7
Aroclor-1260-1	6.188	6.096	6.296	446.310	500.000	-10.7
Aroclor-1260-2	6.378	6.284	6.484	423.140	500.000	-15.4
Aroclor-1260-3	6.528	6.435	6.635	414.840	500.000	-17.0
Aroclor-1260-4	6.997	6.903	7.103	429.320	500.000	-14.1
Aroclor-1260-5	7.239	7.146	7.346	419.690	500.000	-16.1
Decachlorobiphenyl	8.607	8.514	8.714	51.340	50.000	2.7
Tetrachloro-m-xylene	3.648	3.551	3.751	49.130	50.000	-1.7