

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075352.D **Time Analyzed:** 11:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.813	5.708	5.908	512.700	500.000	2.5
Aroclor-1016-2	5.834	5.729	5.929	511.450	500.000	2.3
Aroclor-1016-3	5.897	5.792	5.992	505.880	500.000	1.2
Aroclor-1016-4	5.994	5.889	6.089	539.950	500.000	8.0
Aroclor-1016-5	6.287	6.181	6.381	536.140	500.000	7.2
Aroclor-1260-1	7.405	7.299	7.499	524.760	500.000	5.0
Aroclor-1260-2	7.658	7.551	7.751	496.220	500.000	-0.8
Aroclor-1260-3	8.016	7.909	8.109	551.580	500.000	10.3
Aroclor-1260-4	8.244	8.137	8.337	583.740	500.000	16.7
Aroclor-1260-5	8.570	8.464	8.664	568.790	500.000	13.8
Decachlorobiphenyl	10.445	10.334	10.534	51.500	50.000	3.0
Tetrachloro-m-xylene	4.660	4.555	4.755	53.190	50.000	6.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075352.D **Time Analyzed:** 11:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.892	4.793	4.993	516.440	500.000	3.3
Aroclor-1016-2	4.951	4.851	5.051	517.110	500.000	3.4
Aroclor-1016-3	5.070	4.971	5.171	468.630	500.000	-6.3
Aroclor-1016-4	5.112	5.011	5.211	494.800	500.000	-1.0
Aroclor-1016-5	5.325	5.225	5.425	501.970	500.000	0.4
Aroclor-1260-1	6.354	6.255	6.455	496.790	500.000	-0.6
Aroclor-1260-2	6.540	6.441	6.641	492.500	500.000	-1.5
Aroclor-1260-3	6.695	6.595	6.795	483.140	500.000	-3.4
Aroclor-1260-4	7.164	7.065	7.265	451.220	500.000	-9.8
Aroclor-1260-5	7.404	7.304	7.504	484.470	500.000	-3.1
Decachlorobiphenyl	8.804	8.703	8.903	59.670	50.000	19.3
Tetrachloro-m-xylene	3.796	3.696	3.896	54.460	50.000	8.9