

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075367.D **Time Analyzed:** 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.813	5.708	5.908	509.260	500.000	1.9
Aroclor-1016-2	5.834	5.729	5.929	506.180	500.000	1.2
Aroclor-1016-3	5.897	5.792	5.992	503.020	500.000	0.6
Aroclor-1016-4	5.994	5.889	6.089	532.270	500.000	6.5
Aroclor-1016-5	6.286	6.181	6.381	527.390	500.000	5.5
Aroclor-1260-1	7.404	7.299	7.499	492.460	500.000	-1.5
Aroclor-1260-2	7.657	7.551	7.751	473.360	500.000	-5.3
Aroclor-1260-3	8.015	7.909	8.109	515.440	500.000	3.1
Aroclor-1260-4	8.243	8.137	8.337	529.850	500.000	6.0
Aroclor-1260-5	8.570	8.464	8.664	535.970	500.000	7.2
Decachlorobiphenyl	10.444	10.334	10.534	50.160	50.000	0.3
Tetrachloro-m-xylene	4.660	4.555	4.755	54.060	50.000	8.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/26/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075367.D **Time Analyzed:** 16:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.895	4.793	4.993	545.640	500.000	9.1
Aroclor-1016-2	4.953	4.851	5.051	547.010	500.000	9.4
Aroclor-1016-3	5.072	4.971	5.171	507.650	500.000	1.5
Aroclor-1016-4	5.114	5.011	5.211	521.840	500.000	4.4
Aroclor-1016-5	5.328	5.225	5.425	500.810	500.000	0.2
Aroclor-1260-1	6.356	6.255	6.455	428.080	500.000	-14.4
Aroclor-1260-2	6.542	6.441	6.641	434.700	500.000	-13.1
Aroclor-1260-3	6.696	6.595	6.795	419.470	500.000	-16.1
Aroclor-1260-4	7.166	7.065	7.265	375.600	500.000	-24.9
Aroclor-1260-5	7.405	7.304	7.504	412.990	500.000	-17.4
Decachlorobiphenyl	8.805	8.703	8.903	52.780	50.000	5.6
Tetrachloro-m-xylene	3.799	3.696	3.896	54.320	50.000	8.6