

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3181

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/09/2025 09/09/2025

Continuing Calib Time: 17:38

Initial Calibration Time(s): 11:32 13:08

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.28	7.29	7.19	7.39	0.01
2,4-D	8.41	8.41	8.31	8.51	0.00
2,4,5-TP(Silvex)	9.28	9.29	9.19	9.39	0.01

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: ENTA05

Lab Code: ACE

SDG NO.: Q3181

Continuing Calib Date: 09/29/2025

Initial Calibration Date(s): 09/09/2025 09/09/2025

Continuing Calib Time: 17:38

Initial Calibration Time(s): 11:32 13:08

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.74	7.74	7.64	7.84	0.00
2,4-D	8.99	9.00	8.90	9.10	0.01
2,4,5-TP(Silvex)	9.89	9.89	9.79	9.99	0.00