

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031867.D **Time Analyzed:** 17:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.283	9.185	9.385	730.800	712.500	2.6
2,4-D	8.408	8.314	8.514	716.740	705.000	1.7
2,4-DCAA	7.282	7.185	7.385	752.720	750.000	0.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL01 **Date Analyzed:** 09/29/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031867.D **Time Analyzed:** 17:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.886	9.793	9.993	693.680	712.500	-2.6
2,4-D	8.992	8.900	9.100	688.160	705.000	-2.4
2,4-DCAA	7.736	7.643	7.843	723.550	750.000	-3.5