

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031879.D **Time Analyzed:** 22:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.281	9.185	9.385	722.810	712.500	1.4
2,4-D	8.408	8.314	8.514	720.050	705.000	2.1
2,4-DCAA	7.281	7.185	7.385	751.810	750.000	0.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/09/2025 09/09/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS031879.D **Time Analyzed:** 22:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.890	9.793	9.993	694.770	712.500	-2.5
2,4-D	8.997	8.900	9.100	689.640	705.000	-2.2
2,4-DCAA	7.742	7.643	7.843	740.780	750.000	-1.2