

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092925\
 Data File : PS031872.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Sep 2025 19:43
 Operator : AR\AJ
 Sample : Q3181-03MS
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 WC-A3-01-CMS

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 09/30/2025
 Supervised By :mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Sep 30 02:28:18 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
 Quant Title : 8080.M
 QLast Update : Tue Sep 09 13:56:38 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.280	7.739	1691.1E6	527.8E6	469.562	580.747
Target Compounds							
1) T	Dalapon	2.664	2.680	615.1E6	1066.6E6	106.405m	378.559m#
2) T	3,5-DICHL...	6.448	6.695	1262.4E6	335.9E6	236.694	220.258
5) T	DICAMBA	7.465	7.937	3856.7E6	1701.8E6	247.279	261.543
6) T	MCP	7.641	8.033	186.8E6	27708384	20.900	14.712 #
7) T	MCPA	7.790	8.278	275.5E6	53845183	24.792	19.222
8) T	DICHLORPROP	8.172	8.655	1045.4E6	407.4E6	293.519	279.958
9) T	2,4-D	8.407	8.997	1150.7E6	541.5E6	335.900	329.169
10) T	Pentachlo...	8.703	9.506	9541.3E6	6294.2E6	179.707	158.583
11) T	2,4,5-TP ...	9.280	9.889	6280.9E6	6751.3E6	299.464	448.462 #
12) T	2,4,5-T	9.575	10.319	5504.4E6	3818.3E6	300.816	283.739
13) T	2,4-DB	10.150	10.886	663.2E6	295.5E6	284.303	276.122
14) T	DINOSEB	11.348	11.263	1920.8E6	1145.8E6	133.179	117.868
15) T	Picloram	11.174	12.382	4687.9E6	4831.0E6	260.281	203.982m
16) T	DCPA	11.647	12.304	7472.5E6	6291.2E6	277.275	261.001m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS092925\
Data File : PS031872.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Sep 2025 19:43
Operator : AR\AJ
Sample : Q3181-03MS
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
WC-A3-01-CMS

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 09/30/2025
Supervised By : mohammad ahmed 10/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Sep 30 02:28:18 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS090925.M
Quant Title : 8080.M
QLast Update : Tue Sep 09 13:56:38 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

