

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090445.D **Time Analyzed:** 23:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.063	8.968	9.168	51.010	50.000	2.0
Endrin	6.568	6.469	6.669	59.240	50.000	18.5
gamma-BHC (Lindane)	4.325	4.226	4.426	57.270	50.000	14.5
Heptachlor	4.923	4.825	5.025	58.150	50.000	16.3
Heptachlor epoxide	5.684	5.586	5.786	56.760	50.000	13.5
Methoxychlor	7.487	7.390	7.590	56.670	50.000	13.3
Tetrachloro-m-xylene	3.545	3.446	3.646	50.800	50.000	1.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL02 **Date Analyzed:** 09/29/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090445.D **Time Analyzed:** 23:19

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.061	7.964	8.164	49.880	50.000	-0.2
Endrin	5.781	5.682	5.882	54.490	50.000	9.0
gamma-BHC (Lindane)	3.723	3.624	3.824	53.560	50.000	7.1
Heptachlor	4.075	3.976	4.176	54.030	50.000	8.1
Heptachlor epoxide	4.864	4.766	4.966	54.300	50.000	8.6
Methoxychlor	6.744	6.646	6.846	54.210	50.000	8.4
Tetrachloro-m-xylene	2.875	2.776	2.976	49.580	50.000	-0.8