

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/30/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090449.D **Time Analyzed:** 11:00

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.074	8.968	9.168	47.940	50.000	-4.1
Endrin	6.576	6.469	6.669	55.090	50.000	10.2
gamma-BHC (Lindane)	4.333	4.226	4.426	51.160	50.000	2.3
Heptachlor	4.931	4.825	5.025	52.930	50.000	5.9
Heptachlor epoxide	5.692	5.586	5.786	52.610	50.000	5.2
Methoxychlor	7.496	7.390	7.590	55.510	50.000	11.0
Tetrachloro-m-xylene	3.552	3.446	3.646	45.550	50.000	-8.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL03 **Date Analyzed:** 09/30/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090449.D **Time Analyzed:** 11:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	8.064	7.964	8.164	47.210	50.000	-5.6
Endrin	5.782	5.682	5.882	50.380	50.000	0.8
gamma-BHC (Lindane)	3.723	3.624	3.824	48.390	50.000	-3.2
Heptachlor	4.076	3.976	4.176	48.830	50.000	-2.3
Heptachlor epoxide	4.865	4.766	4.966	48.910	50.000	-2.2
Methoxychlor	6.746	6.646	6.846	50.260	50.000	0.5
Tetrachloro-m-xylene	2.874	2.776	2.976	44.920	50.000	-10.2