

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090463.D **Time Analyzed:** 17:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.065	8.968	9.168	46.930	50.000	-6.1
Endrin	6.568	6.469	6.669	54.950	50.000	9.9
gamma-BHC (Lindane)	4.326	4.226	4.426	52.610	50.000	5.2
Heptachlor	4.924	4.825	5.025	53.650	50.000	7.3
Heptachlor epoxide	5.685	5.586	5.786	53.830	50.000	7.7
Methoxychlor	7.488	7.390	7.590	52.970	50.000	5.9
Tetrachloro-m-xylene	3.545	3.446	3.646	47.080	50.000	-5.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ENTA05
Lab Code: ACE **SDG NO.:** Q3181
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/18/2025 09/18/2025

Client Sample No.: CCAL04 **Date Analyzed:** 09/30/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD090463.D **Time Analyzed:** 17:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	8.061	7.964	8.164	43.850	50.000	-12.3
Endrin	5.781	5.682	5.882	50.970	50.000	1.9
gamma-BHC (Lindane)	3.723	3.624	3.824	49.090	50.000	-1.8
Heptachlor	4.075	3.976	4.176	49.880	50.000	-0.2
Heptachlor epoxide	4.864	4.766	4.966	48.960	50.000	-2.1
Methoxychlor	6.745	6.646	6.846	49.180	50.000	-1.6
Tetrachloro-m-xylene	2.876	2.776	2.976	45.660	50.000	-8.7