

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD091825

Review By	Abdul	Review On	10/2/2025 10:15:46 AM
Supervise By	Sohil	Supervise On	10/2/2025 10:03:44 AM
SubDirectory	PD091825	HP Acquire Method	HP Processing Method pd091825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090326.D	18 Sep 2025 09:36		AR\AJ	Ok
2	I.BLK	I.BLK	PD090327.D	18 Sep 2025 09:49		AR\AJ	Ok
3	PEM	PEM	PD090328.D	18 Sep 2025 10:03		AR\AJ	Ok,M
4	RESCHK	RESCHK	PD090329.D	18 Sep 2025 10:16		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PD090330.D	18 Sep 2025 10:30		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PD090331.D	18 Sep 2025 10:43		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PD090332.D	18 Sep 2025 10:57		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PD090333.D	18 Sep 2025 11:11		AR\AJ	Ok
9	PSTDICC005	PSTDICC005	PD090334.D	18 Sep 2025 11:24		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PD090335.D	18 Sep 2025 11:38		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PD090336.D	18 Sep 2025 11:51		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PD090337.D	18 Sep 2025 12:05		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PD090338.D	18 Sep 2025 12:18		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PD090339.D	18 Sep 2025 12:32		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PD090340.D	18 Sep 2025 12:45		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PD090341.D	18 Sep 2025 12:59		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PD090342.D	18 Sep 2025 13:12		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PD090343.D	18 Sep 2025 13:26		AR\AJ	Ok

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD091825

Review By	Abdul	Review On	10/2/2025 10:15:46 AM
Supervise By	Sohil	Supervise On	10/2/2025 10:03:44 AM
SubDirectory	PD091825	HP Acquire Method	HP Processing Method pd091825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PD090344.D	18 Sep 2025 13:40		AR\AJ	Ok
20	PSTDICV050	ICVPD091825	PD090345.D	18 Sep 2025 13:53		AR\AJ	Ok,M
21	PCHLORICV500	ICVPD091825CHLOR	PD090346.D	18 Sep 2025 14:07		AR\AJ	Ok
22	PTOXICV500	ICVPD091825TOX	PD090347.D	18 Sep 2025 15:21		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD092925

Review By	Abdul	Review On	9/30/2025 8:42:12 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:41:12 AM
SubDirectory	PD092925	HP Acquire Method	HP Processing Method PD091825

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090415.D	29 Sep 2025 10:45		AR\AJ	Ok
2	I.BLK	I.BLK	PD090416.D	29 Sep 2025 11:15		AR\AJ	Ok,M
3	PEM	PEM	PD090417.D	29 Sep 2025 11:29		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD090418.D	29 Sep 2025 14:24		AR\AJ	Ok,M
5	PB169830BL	PB169830BL	PD090419.D	29 Sep 2025 14:44		AR\AJ	Ok
6	PB169830BS	PB169830BS	PD090420.D	29 Sep 2025 15:59		AR\AJ	Ok,M
7	PB169830BSD	PB169830BSD	PD090421.D	29 Sep 2025 16:20	Methoxychlor recovery fail	AR\AJ	Not Ok
8	PB169882BL	PB169882BL	PD090422.D	29 Sep 2025 17:24		AR\AJ	Ok
9	PB169882BS	PB169882BS	PD090423.D	29 Sep 2025 17:37		AR\AJ	Ok,M
10	PB169882TB	PB169882TB	PD090424.D	29 Sep 2025 17:51		AR\AJ	Ok
11	Q3177-01	EME-TS12-01	PD090425.D	29 Sep 2025 18:05		AR\AJ	Ok
12	Q3177-02	EME-TS13-01	PD090426.D	29 Sep 2025 18:18		AR\AJ	Ok,M
13	Q3177-03	EME-TS14-02	PD090427.D	29 Sep 2025 18:32		AR\AJ	Ok
14	Q3179-01	EME-TS14-01	PD090428.D	29 Sep 2025 18:46		AR\AJ	Ok,M
15	Q3179-02MS	EME-TS14-01MS	PD090429.D	29 Sep 2025 18:59		AR\AJ	Ok,M
16	Q3179-03MSD	EME-TS14-01MSD	PD090430.D	29 Sep 2025 19:13		AR\AJ	Ok,M
17	I.BLK	I.BLK	PD090431.D	29 Sep 2025 19:27		AR\AJ	Ok
18	PEM	PEM	PD090432.D	29 Sep 2025 19:40		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD092925

Review By	Abdul	Review On	9/30/2025 8:42:12 AM
Supervise By	mohammad	Supervise On	10/1/2025 1:41:12 AM
SubDirectory	PD092925	HP Acquire Method	HP Processing Method PD091825
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PSTDCCC050	PSTDCCC050	PD090433.D	29 Sep 2025 19:54		AR\AJ	Ok,M
20	PB169874BL	PB169874BL	PD090434.D	29 Sep 2025 20:35		AR\AJ	Ok
21	PB169874BS	PB169874BS	PD090435.D	29 Sep 2025 20:49	Methoxychlor recovery fail	AR\AJ	Not Ok
22	PB169840TB	PB169840TB	PD090436.D	29 Sep 2025 21:02		AR\AJ	Ok
23	Q3181-03	WC-A3-01-C	PD090437.D	29 Sep 2025 21:16		AR\AJ	Ok,M
24	Q3181-03MS	WC-A3-01-CMS	PD090438.D	29 Sep 2025 21:30	Heptachlor epoxide recovery fail	AR\AJ	Ok,M
25	Q3181-03MSD	WC-A3-01-CMSD	PD090439.D	29 Sep 2025 21:43	Heptachlor epoxide recovery fail	AR\AJ	Ok,M
26	Q3181-07	WC-A3-02-C	PD090440.D	29 Sep 2025 21:57		AR\AJ	Ok,M
27	Q3181-11	WC-A3-03-C	PD090441.D	29 Sep 2025 22:11		AR\AJ	Ok,M
28	Q3208-06	SU-ESM-COMP-01	PD090442.D	29 Sep 2025 22:25		AR\AJ	Ok,M
29	Q3208-12	SU-ESM-COMP-02	PD090443.D	29 Sep 2025 22:38		AR\AJ	Ok,M
30	I.BLK	I.BLK	PD090444.D	29 Sep 2025 22:52		AR\AJ	Ok
31	PSTDCCC050	PSTDCCC050	PD090445.D	29 Sep 2025 23:19		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD093025

Review By	Abdul	Review On	10/1/2025 8:38:03 AM
Supervise By	mohammad	Supervise On	10/3/2025 1:30:28 AM
SubDirectory	PD093025	HP Acquire Method	HP Processing Method pd091825 8081

STD. NAME	STD REF.#
Tune/Reschk	PP24890,PP24651
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,P24763,PP24764
CCC	PP24751,PP24756,PP24761
Internal Standard/PEM	
ICV/I.BLK	PP24754,PP24759,PP24761
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PD090446.D	30 Sep 2025 09:52		AR\AJ	Ok
2	I.BLK	I.BLK	PD090447.D	30 Sep 2025 10:06		AR\AJ	Ok
3	PEM	PEM	PD090448.D	30 Sep 2025 10:19		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PD090449.D	30 Sep 2025 11:00		AR\AJ	Ok,M
5	Q3200-01	ETGI-367	PD090450.D	30 Sep 2025 12:59		AR\AJ	Ok,M
6	Q3206-01	TR-06-092525	PD090451.D	30 Sep 2025 13:12		AR\AJ	Ok,M
7	Q3208-01	SU-ESM-COMP-01	PD090452.D	30 Sep 2025 13:26		AR\AJ	Ok,M
8	Q3208-07	SU-ESM-COMP-02	PD090453.D	30 Sep 2025 13:39		AR\AJ	Ok,M
9	Q3210-01	SG-1	PD090454.D	30 Sep 2025 13:53		AR\AJ	Ok,M
10	Q3200-01MS	ETGI-367MS	PD090455.D	30 Sep 2025 14:07	Methoxychlor recovery fail	AR\AJ	Ok,M
11	Q3200-01MSD	ETGI-367MSD	PD090456.D	30 Sep 2025 14:20		AR\AJ	Ok,M
12	PB169857BL	PB169857BL	PD090457.D	30 Sep 2025 14:49		AR\AJ	Ok
13	PB169857BS	PB169857BS	PD090458.D	30 Sep 2025 15:37	DCB high in 1st column	AR\AJ	Not Ok
14	PB169830BS	PB169830BS	PD090459.D	30 Sep 2025 16:09		AR\AJ	Ok,M
15	PB169830BSD	PB169830BSD	PD090460.D	30 Sep 2025 16:26		AR\AJ	Ok,M
16	PB169874BS	PB169874BS	PD090461.D	30 Sep 2025 16:39		AR\AJ	Ok,M
17	I.BLK	I.BLK	PD090462.D	30 Sep 2025 16:55		AR\AJ	Ok
18	PSTDCCC050	PSTDCCC050	PD090463.D	30 Sep 2025 17:09		AR\AJ	Ok,M

Instrument ID: ECD_D

Daily Analysis Runlog For Sequence/QC Batch ID # PD093025

Review By	Abdul	Review On	10/1/2025 8:38:03 AM
Supervise By	mohammad	Supervise On	10/3/2025 1:30:28 AM
SubDirectory	PD093025	HP Acquire Method	HP Processing Method pd091825 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP24890,PP24651		
Initial Calibration Stds	PP24744,PP24746,PP24748,PP24750,PP24751,PP24752,PP24753,PP24755,PP24756,PP24757,PP24758,PP24759,PP24760,PP24761,PP24762,PP24763,PP24764		
CCC	PP24751,PP24756,PP24761		
Internal Standard/PEM			
ICV/I.BLK	PP24754,PP24759,PP24761		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB169902BL	PB169902BL	PD090464.D	30 Sep 2025 17:54		AR\AJ	Ok
20	PB169902BS	PB169902BS	PD090465.D	30 Sep 2025 18:08	Methoxychlor recovery fail	AR\AJ	Not Ok
21	PB169870TB	PB169870TB	PD090466.D	30 Sep 2025 18:22		AR\AJ	Ok
22	Q3231-04	WC-1	PD090467.D	30 Sep 2025 18:35		AR\AJ	Ok
23	Q3231-04MS	WC-1MS	PD090468.D	30 Sep 2025 18:49		AR\AJ	Ok,M
24	Q3231-04MSD	WC-1MSD	PD090469.D	30 Sep 2025 19:02		AR\AJ	Ok,M
25	Q3231-08	WC-2	PD090470.D	30 Sep 2025 19:16		AR\AJ	Ok
26	Q3231-12	WC-3	PD090471.D	30 Sep 2025 19:30		AR\AJ	Ok
27	Q3231-16	WC-4	PD090472.D	30 Sep 2025 19:43		AR\AJ	Ok
28	I.BLK	I.BLK	PD090473.D	30 Sep 2025 19:57		AR\AJ	Ok
29	PSTDCCC050	PSTDCCC050	PD090474.D	30 Sep 2025 21:19		AR\AJ	Ok,M

M : Manual Integration