284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT	Date Collected:	09/24/25
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	09/24/25
Client Sample ID:	WC-A3-01-CMS	SDG No.:	Q3181
Lab Sample ID:	Q3181-03MS	Matrix:	TCLP
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	100 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	TCLP BNA
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG064463.D	1	09/26/25 11:10	09/29/25 17:15	PB169873

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
110-86-1	Pyridine	0.28		0.013	0.050	mg/L
106-46-7	1,4-Dichlorobenzene	0.36		0.0053	0.050	mg/L
95-48-7	2-Methylphenol	0.38		0.011	0.050	mg/L
65794-96-9	3+4-Methylphenols	0.37		0.011	0.10	mg/L
67-72-1	Hexachloroethane	0.35		0.0065	0.050	mg/L
98-95-3	Nitrobenzene	0.40		0.0076	0.050	mg/L
87-68-3	Hexachlorobutadiene	0.36		0.0054	0.050	mg/L
88-06-2	2,4,6-Trichlorophenol	0.45		0.0051	0.050	mg/L
95-95-4	2,4,5-Trichlorophenol	0.47		0.0062	0.050	mg/L
121-14-2	2,4-Dinitrotoluene	0.49		0.012	0.050	mg/L
118-74-1	Hexachlorobenzene	0.50		0.0052	0.050	mg/L
87-86-5	Pentachlorophenol	1.00	E	0.016	0.10	mg/L
SURROGATES						
367-12-4	2-Fluorophenol	120		15 (23) - 110 (138)	80%	SPK: 150
13127-88-3	Phenol-d6	109		15 (10) - 110 (134)	73%	SPK: 150
4165-60-0	Nitrobenzene-d5	79.5		30 (67) - 130 (132)	80%	SPK: 100
321-60-8	2-Fluorobiphenyl	81.4		30 (52) - 130 (132)	81%	SPK: 100
118-79-6	2,4,6-Tribromophenol	144		15 (44) - 110 (137)	96%	SPK: 150
1718-51-0	Terphenyl-d14	92.4		30 (42) - 130 (152)	92%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	20000	7.836			
1146-65-2	Naphthalene-d8	82000	10.621			
15067-26-2	Acenaphthene-d10	58700	14.463			
1517-22-2	Phenanthrene-d10	139000	17.201			
1719-03-5	Chrysene-d12	157000	21.431			
1520-96-3	Perylene-d12	170000	24.422			

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	ENTACT			Date Collected:	09/24/25
Project:	540 Degraw St, Brooklyn, NY - E9309			Date Received:	09/24/25
Client Sample ID:	WC-A3-01-CMS			SDG No.:	Q3181
Lab Sample ID:	Q3181-03MS			Matrix:	TCLP
Analytical Method:	8270E			% Solid:	0
Sample Wt/Vol:	100	Units:	mL	Final Vol:	1000 uL
Soil Aliquot Vol:			uL	Test:	TCLP BNA
Extraction Type :			Decanted : N	Level :	LOW
Injection Volume :			GPC Factor : 1.0	GPC Cleanup : N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BG064463.D	1	09/26/25 11:10	09/29/25 17:15	PB169873

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products