

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG092525

Review By	Rahul	Review On	9/26/2025 9:06:10 AM
Supervise By		Supervise On	
SubDirectory	BG092525	HP Acquire Method	BNA_G
		HP Processing Method	BG092525
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13175,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064440.D	25 Sep 2025 11:05	RC/JU	Ok
2	SSTDICC2.5	BG064441.D	25 Sep 2025 11:55	RC/JU	Ok
3	SSTDICC005	BG064442.D	25 Sep 2025 12:37	RC/JU	Ok,NR
4	SSTDICC010	BG064443.D	25 Sep 2025 13:18	RC/JU	Not Ok
5	SSTDICC020	BG064444.D	25 Sep 2025 13:59	RC/JU	Ok,NR
6	SSTDCCC040	BG064445.D	25 Sep 2025 14:41	RC/JU	Not Ok
7	SSTDICC050	BG064446.D	25 Sep 2025 15:22	RC/JU	Ok,NR
8	SSTDCC060	BG064447.D	25 Sep 2025 16:03	RC/JU	Not Ok
9	SSTDICC080	BG064448.D	25 Sep 2025 16:44	RC/JU	Ok,NR
10	SSTDICC010	BG064449.D	25 Sep 2025 18:04	RC/JU	Ok,NR
11	SSTDICCC040	BG064450.D	25 Sep 2025 18:45	RC/JU	Ok,NR
12	SSTDICC060	BG064451.D	25 Sep 2025 19:26	RC/JU	Ok,NR
13	SSTDICV040	BG064452.D	25 Sep 2025 20:49	RC/JU	Ok,NR
14	PB169726BL	BG064453.D	25 Sep 2025 21:30	RC/JU	Ok
15	PB169726BS	BG064454.D	25 Sep 2025 22:11	RC/JU	Ok,NR

M : Manual Integration

Instrument ID: **BNA_G**

Daily Analysis Runlog For Sequence/QC Batch ID # BG092925

Review By	rahul	Review On	9/30/2025 8:41:59 AM
Supervise By	Jagrut	Supervise On	9/30/2025 11:09:02 AM
SubDirectory	BG092925	HP Acquire Method	BNA_G
		HP Processing Method	BG092525
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6857,SP6858,SP6859,SP6860,SP6861,SP6862,SP6863,SP6864		
CCC	SP6860		
Internal Standard/PEM	S13176,10ul/1000ul sample		
ICV/I.BLK	SP6872		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BG064455.D	29 Sep 2025 11:41	RC/JU	Ok
2	SSTDCCC040	BG064456.D	29 Sep 2025 12:21	RC/JU	Ok,M
3	PB169873BL	BG064457.D	29 Sep 2025 13:02	RC/JU	Ok
4	PB169873BS	BG064458.D	29 Sep 2025 13:43	RC/JU	Ok,M
5	Q3183-10	BG064459.D	29 Sep 2025 14:31	RC/JU	Ok
6	Q3183-11	BG064460.D	29 Sep 2025 15:12	RC/JU	Ok
7	Q3208-12	BG064461.D	29 Sep 2025 15:53	RC/JU	Ok
8	Q3181-03	BG064462.D	29 Sep 2025 16:34	RC/JU	Ok,M
9	Q3181-03MS	BG064463.D	29 Sep 2025 17:15	RC/JU	Ok,M
10	Q3181-03MSD	BG064464.D	29 Sep 2025 17:56	RC/JU	Ok,M
11	Q3181-07	BG064465.D	29 Sep 2025 18:37	RC/JU	Ok
12	Q3181-11	BG064466.D	29 Sep 2025 19:19	RC/JU	Ok
13	Q3208-06	BG064467.D	29 Sep 2025 20:00	RC/JU	Ok
14	Q3232-01	BG064468.D	29 Sep 2025 20:41	RC/JU	Ok
15	Q3232-01DL	BG064469.D	29 Sep 2025 21:22	RC/JU	Not Ok
16	Q3215-01	BG064470.D	29 Sep 2025 22:03	RC/JU	Ok,M
17	Q3215-01DL	BG064471.D	29 Sep 2025 22:44	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP091225

Review By	Rahul	Review On	9/12/2025 10:23:37 AM
Supervise By	Jagrut	Supervise On	9/12/2025 11:49:47 AM
SubDirectory	BP091225	HP Acquire Method	BNA_P
		HP Processing Method	BP091225
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840		
CCC	SP6836		
Internal Standard/PEM	S13174,10ul/1000ul sample		
ICV/I.BLK	SP6796		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP025724.D	11 Sep 2025 08:09	CG/JU	Ok
2	SSTDICC2.5	BP025725.D	11 Sep 2025 08:56	CG/JU	Ok,M
3	SSTDICC005	BP025726.D	11 Sep 2025 09:37	CG/JU	Ok,M
4	SSTDICC010	BP025727.D	11 Sep 2025 10:18	CG/JU	Ok,M
5	SSTDICC020	BP025728.D	11 Sep 2025 10:59	CG/JU	Ok,M
6	SSTDICCC040	BP025729.D	11 Sep 2025 12:21	CG/JU	Ok,M
7	SSTDICC050	BP025730.D	11 Sep 2025 13:02	CG/JU	Not Ok
8	SSTDICC060	BP025731.D	11 Sep 2025 13:43	CG/JU	Ok
9	SSTDICC080	BP025732.D	11 Sep 2025 14:24	CG/JU	Ok
10	SSTDICC050	BP025733.D	11 Sep 2025 15:20	CG/JU	Ok,M
11	SSTDICV040	BP025734.D	11 Sep 2025 16:54	CG/JU	Ok,M
12	PB169633BL	BP025735.D	11 Sep 2025 17:35	CG/JU	Ok
13	DFTPP	BP025736.D	12 Sep 2025 09:03	CG/JU	Ok
14	SSTDCCC040	BP025737.D	12 Sep 2025 10:24	CG/JU	Ok
15	PB169490BL	BP025738.D	12 Sep 2025 11:05	CG/JU	Ok
16	Q2893-02	BP025739.D	12 Sep 2025 11:46	CG/JU	Ok,M
17	Q2893-08	BP025740.D	12 Sep 2025 12:27	CG/JU	Ok,M
18	Q2893-01	BP025741.D	12 Sep 2025 13:08	CG/JU	Ok,M
19	Q2893-07	BP025742.D	12 Sep 2025 13:49	CG/JU	Ok,M
20	PB169499BL	BP025743.D	12 Sep 2025 14:30	CG/JU	Ok
21	SSTDCCC040	BP025744.D	12 Sep 2025 15:35	CG/JU	Ok,M

M : Manual Integration

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP092925

Review By	rahul	Review On	9/30/2025 8:35:30 AM
Supervise By	Jagrut	Supervise On	9/30/2025 11:08:17 AM
SubDirectory	BP092925	HP Acquire Method	BNA_P
		HP Processing Method	BP091225
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840		
CCC	SP6836		
Internal Standard/PEM	S13176,10ul/1000ul sample		
ICV/I.BLK	SP6796		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP025874.D	29 Sep 2025 10:11	CG/JU	Ok
2	SSTDCCC040	BP025875.D	29 Sep 2025 10:51	CG/JU	Ok
3	PB169840TB	BP025876.D	29 Sep 2025 11:32	CG/JU	Ok
4	Q3183-12	BP025877.D	29 Sep 2025 12:20	CG/JU	Ok
5	Q3200-01MS	BP025878.D	29 Sep 2025 13:01	CG/JU	Ok
6	Q3200-01MSD	BP025879.D	29 Sep 2025 13:42	CG/JU	Ok,M
7	Q3210-01	BP025880.D	29 Sep 2025 14:23	CG/JU	Ok
8	Q3209-01	BP025881.D	29 Sep 2025 15:04	CG/JU	Ok
9	Q3208-07	BP025882.D	29 Sep 2025 15:45	CG/JU	Ok,M
10	Q3206-01	BP025883.D	29 Sep 2025 16:26	CG/JU	Ok,M
11	Q3208-01	BP025884.D	29 Sep 2025 17:07	CG/JU	Ok,M
12	Q3212-01	BP025885.D	29 Sep 2025 17:49	CG/JU	Ok,M
13	Q3231-01	BP025886.D	29 Sep 2025 18:30	CG/JU	Ok,M
14	Q3220-01	BP025887.D	29 Sep 2025 19:11	CG/JU	Ok,M
15	Q3221-01	BP025888.D	29 Sep 2025 19:52	CG/JU	Ok,M
16	Q3222-01	BP025889.D	29 Sep 2025 20:33	CG/JU	Ok,M
17	Q3231-09	BP025890.D	29 Sep 2025 21:15	CG/JU	Ok

M : Manual Integration