

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3177

Analytical Method: SW8270E

Client: Weston Solutions, Inc.

DataFile: BG064463.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q3181-03MS		Client Sample ID: WC-A3-01-CMS									
Pyridine	500	0	280	ug/L	56				20 (10)	160 (109)	
1,4-Dichlorobenzene	500	0	360	ug/L	72				70 (55)	130 (125)	
2-Methylphenol	500	0	380	ug/L	76				70 (60)	130 (131)	
3+4-Methylphenols	500	0	370	ug/L	74				20 (54)	160 (136)	
Hexachloroethane	500	0	350	ug/L	70				20 (19)	160 (146)	
Nitrobenzene	500	0	400	ug/L	80				70 (62)	130 (112)	
Hexachlorobutadiene	500	0	360	ug/L	72				70 (52)	130 (125)	
2,4,6-Trichlorophenol	500	0	450	ug/L	90				70 (78)	130 (112)	
2,4,5-Trichlorophenol	500	0	470	ug/L	94				70 (71)	130 (111)	
2,4-Dinitrotoluene	500	0	490	ug/L	98				70 (74)	130 (137)	
Hexachlorobenzene	500	0	500	ug/L	100				70 (72)	130 (115)	
Pentachlorophenol	1000	0	1000	ug/L	100				20 (52)	160 (162)	