

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3177

Analytical Method: 8270E

Client: Weston Solutions, Inc.

DataFile: BG064458.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB169873BS	Pyridine	50	33.6	ug/L	67				20 (29)	160 (97)	
	1,4-Dichlorobenzene	50	42.8	ug/L	86				70 (76)	130 (103)	
	2-Methylphenol	50	45.3	ug/L	91				70 (69)	130 (109)	
	3+4-Methylphenols	50	44.6	ug/L	89				20 (67)	160 (106)	
	Hexachloroethane	50	40.5	ug/L	81				20 (76)	160 (118)	
	Nitrobenzene	50	47.2	ug/L	94				70 (58)	130 (106)	
	Hexachlorobutadiene	50	44.3	ug/L	89				70 (69)	130 (101)	
	2,4,6-Trichlorophenol	50	50.7	ug/L	101				70 (61)	130 (110)	
	2,4,5-Trichlorophenol	50	50.7	ug/L	101				70 (70)	130 (106)	
	2,4-Dinitrotoluene	50	51.2	ug/L	102				70 (60)	130 (115)	
	Hexachlorobenzene	50	44.5	ug/L	89				70 (73)	130 (106)	
	Pentachlorophenol	100	97.0	ug/L	97				20 (47)	160 (114)	