

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3177

Client ID : SSTDCCC040

Date Analyzed: 09/29/2025

Lab File ID: BP025875.D

Time Analyzed: 10:51

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	234395	7.743	868295	10.51	527881	14.37
UPPER LIMIT	468790	8.243	1736590	11.013	1055760	14.872
LOWER LIMIT	117198	7.243	434148	10.013	263941	13.872
EPA SAMPLE NO.						
01 PB169840TB	218795	7.74	860397	10.52	587654	14.37

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.