

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3177

Client ID: SSTDCCC040

Date Analyzed: 09/29/2025

Lab File ID: BP025875.D

Time Analyzed: 10:51

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

01

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	977878	17.172	1032520	21.613	1270410	24.965
UPPER LIMIT	1955760	17.672	2065040	22.113	2540820	25.465
LOWER LIMIT	488939	16.672	516260	21.113	635205	24.465
EPA SAMPLE NO.						
PB169840TB	1277550	17.16	1290550	21.61	1166610	24.97

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.