

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN092325

Review By	John Carlone	Review On	9/24/2025 8:10:38 AM
Supervise By	Maresh Dadoda	Supervise On	9/26/2025 1:41:42 AM
SubDirectory	VN092325	HP Acquire Method	HP Processing Method 82N092325W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP135592		
Initial Calibration Stds	VP135674,VP135675,VP135676,VP135677,VP135678		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP135679		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087922.D	23 Sep 2025 08:31		JC\MD	Ok
2	VSTDIC001	VSTDIC001	VN087923.D	23 Sep 2025 08:51	Not used	JC\MD	Not Ok
3	VSTDIC005	VSTDIC005	VN087924.D	23 Sep 2025 09:33		JC\MD	Ok,M
4	VSTDIC020	VSTDIC020	VN087925.D	23 Sep 2025 09:54	Spiked Low	JC\MD	Not Ok
5	VSTDIC050	VSTDIC050	VN087926.D	23 Sep 2025 10:15		JC\MD	Ok,M
6	VSTDIC100	VSTDIC100	VN087927.D	23 Sep 2025 10:36		JC\MD	Ok,M
7	VSTDIC150	VSTDIC150	VN087928.D	23 Sep 2025 10:56		JC\MD	Ok,M
8	IBLK	IBLK	VN087929.D	23 Sep 2025 11:17		JC\MD	Ok
9	VSTDIC020	VSTDIC020	VN087930.D	23 Sep 2025 11:47		JC\MD	Ok,M
10	VSTDICV050	ICVVN092325	VN087931.D	23 Sep 2025 15:08		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN092625

Review By	Semsettin Yesilyurt	Review On	9/29/2025 7:39:58 AM
Supervise By		Supervise On	
SubDirectory	VN092625	HP Acquire Method	HP Processing Method 82N092325W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP135667		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135668,VP135669		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN087960.D	26 Sep 2025 12:11		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN087961.D	26 Sep 2025 12:53	pH#Lot#V12668	JC\MD	Ok,M
3	VN0926MBL01	VN0926MBL01	VN087962.D	26 Sep 2025 13:28		JC\MD	Ok
4	VN0926WBL01	VN0926WBL01	VN087963.D	26 Sep 2025 13:49		JC\MD	Ok
5	VN0926WBS01	VN0926WBS01	VN087964.D	26 Sep 2025 14:10		JC\MD	Ok,M
6	VN0926WBSD01	VN0926WBSD01	VN087965.D	26 Sep 2025 14:45		JC\MD	Ok,M
7	PB169841TB	PB169841TB	VN087966.D	26 Sep 2025 15:06		JC\MD	Ok,M
8	Q3181-01	WC-A3-01-G	VN087967.D	26 Sep 2025 15:27	vial A pH#5.0	JC\MD	Ok
9	Q3181-05	WC-A3-02-G	VN087968.D	26 Sep 2025 15:49	vial A pH#5.0	JC\MD	Ok
10	Q3181-09	WC-A3-03-G	VN087969.D	26 Sep 2025 16:10	vial A pH#5.0	JC\MD	Ok
11	Q3183-10	WC-1	VN087970.D	26 Sep 2025 16:31	vial A pH#5.0	JC\MD	Ok
12	Q3183-11	WC-4	VN087971.D	26 Sep 2025 16:52	vial A pH#5.0	JC\MD	Ok
13	Q3183-12	WC-3	VN087972.D	26 Sep 2025 17:13	vial A pH#5.0	JC\MD	Ok
14	Q3208-06	SU-ESM-COMP-01	VN087973.D	26 Sep 2025 17:34	vial A pH#5.0	JC\MD	Ok
15	Q3208-12	SU-ESM-COMP-02	VN087974.D	26 Sep 2025 17:55	Not purge	JC\MD	Not Ok

M : Manual Integration