

Surrogate Summary

SDG No.: **Q3181**

Client: **ENTACT**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3181-01	WC-A3-01-G	1,2-Dichloroethane-d4	50	59.5	119		70 (74)	130 (125)
		Dibromofluoromethane	50	48.5	97		70 (75)	130 (124)
		Toluene-d8	50	50.4	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.3	97		70 (77)	130 (121)
Q3181-05	WC-A3-02-G	1,2-Dichloroethane-d4	50	55.4	111		70 (74)	130 (125)
		Dibromofluoromethane	50	48.1	96		70 (75)	130 (124)
		Toluene-d8	50	48.0	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.8	90		70 (77)	130 (121)
Q3181-09	WC-A3-03-G	1,2-Dichloroethane-d4	50	57.3	115		70 (74)	130 (125)
		Dibromofluoromethane	50	48.3	97		70 (75)	130 (124)
		Toluene-d8	50	47.8	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.2	90		70 (77)	130 (121)
VN0926WBL01	VN0926WBL01	1,2-Dichloroethane-d4	50	58.4	117		70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101		70 (75)	130 (124)
		Toluene-d8	50	50.3	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.6	89		70 (77)	130 (121)
VN0926WBS01	VN0926WBS01	1,2-Dichloroethane-d4	50	53.6	107		70 (74)	130 (125)
		Dibromofluoromethane	50	50.6	101		70 (75)	130 (124)
		Toluene-d8	50	50.7	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101		70 (77)	130 (121)
VN0926WBSD01	VN0926WBSD01	1,2-Dichloroethane-d4	50	50.9	102		70 (74)	130 (125)
		Dibromofluoromethane	50	48.6	97		70 (75)	130 (124)
		Toluene-d8	50	50.3	101		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.7	97		70 (77)	130 (121)

