

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05  
 Lab Code: ACE SDG NO.: Q3181  
 Lab File ID: VN087961.D Date Analyzed: 09/26/2025  
 Instrument ID: MSVOA\_N Time Analyzed: 12:53  
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 207820        | 8.21  | 397459        | 9.08  | 386821        | 11.85  |
| UPPER LIMIT    | 415640        | 8.706 | 794918        | 9.582 | 773642        | 12.347 |
| LOWER LIMIT    | 103910        | 7.706 | 198730        | 8.582 | 193411        | 11.347 |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| WC-A3-01-G     | 178955        | 8.21  | 392164        | 9.08  | 414447        | 11.85  |
| WC-A3-01-G     | 178955        | 8.21  | 392164        | 9.08  | 414447        | 11.85  |
| WC-A3-02-G     | 183424        | 8.21  | 408536        | 9.08  | 411042        | 11.85  |
| WC-A3-02-G     | 183424        | 8.21  | 408536        | 9.08  | 411042        | 11.85  |
| WC-A3-03-G     | 184198        | 8.21  | 412130        | 9.08  | 421613        | 11.84  |
| WC-A3-03-G     | 184198        | 8.21  | 412130        | 9.08  | 421613        | 11.84  |
| VN0926WBL01    | 239293        | 8.21  | 535067        | 9.08  | 507778        | 11.85  |
| VN0926WBL01    | 239293        | 8.21  | 535067        | 9.08  | 507778        | 11.85  |
| VN0926WBS01    | 204106        | 8.21  | 390365        | 9.08  | 380422        | 11.84  |
| VN0926WBS01    | 204106        | 8.21  | 390365        | 9.08  | 380422        | 11.84  |
| VN0926WBSD01   | 220191        | 8.21  | 400196        | 9.08  | 378862        | 11.85  |
| VN0926WBSD01   | 220191        | 8.21  | 400196        | 9.08  | 378862        | 11.85  |

IS1 = Pentafluorobenzene  
 IS2 = 1,4-Difluorobenzene  
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.