

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: ENTA05
 Lab Code: ACE SDG NO.: Q3181
 Lab File ID: VN087961.D Date Analyzed: 09/26/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	216102	13.77				
UPPER LIMIT	432204	14.27				
LOWER LIMIT	108051	13.27				
EPA SAMPLE NO.						
WC-A3-01-G	187569	13.77				
WC-A3-01-G	187569	13.77				
WC-A3-02-G	193008	13.77				
WC-A3-02-G	193008	13.77				
WC-A3-03-G	196969	13.77				
WC-A3-03-G	196969	13.77				
VN0926WBL01	233948	13.77				
VN0926WBL01	233948	13.77				
VN0926WBS01	204880	13.77				
VN0926WBS01	204880	13.77				
VN0926WBSD01	203685	13.77				
VN0926WBSD01	203685	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.