

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

RunNo.: LB137308

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.98	1	98	90-110	09/25/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.98	1	98	90-110	09/25/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	09/25/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	09/25/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

RunNo.: LB137319

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.494	0.50	99	90-110	09/25/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.502	0.5	100	90-110	09/25/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.505	0.5	101	90-110	09/25/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

RunNo.: LB137320

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.505	0.50	101	90-110	09/25/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.502	0.50	100	90-110	09/25/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.491	0.50	98	90-110	09/25/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

RunNo.: LB137320

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------