

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

Run No.: LB137318

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137318BS							
BOD5	mg/L	198	193		97	1	84.6-115.4	09/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

Run No.: LB137319

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137319BS							
Orthophosphate as P	mg/L	0.5	0.49		98	1	90-110	09/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

Run No.: LB137326

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137326BS							
Oil and Grease	mg/L	20.0	17.1		86	1	78-114	09/26/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

Run No.: LB137328

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137328BS							
TSS	mg/L	550	531		96	1	90-110	09/26/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

Run No.: LB137308

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169820BS							
Ammonia as N	mg/L	1	0.99		99	1	90-110	09/25/2025

Laboratory Control Sample Summary

Client: Holland Manufacturing Co.

SDG No.: Q3184

Project: Pre Treatment Plant 2025

Run No.: LB137320

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB169837BS							
Phosphorus, Total	mg/L	0.50	0.51		102	1	90-110	09/25/2025