

Continuing Calibration Report for SequenceID : FD100125AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049788.D

Aromatic C12-C16	8674705.000	60.000	6.132	8.760	144578.417	135865.661	-6.413
Aromatic C10-C12	4962988.000	40.000	4.388	6.131	124074.700	116056.185	-6.909
Aromatic C16-C21	11282024.000	80.000	8.761	13.050	141025.300	135770.533	-3.870
Aromatic C21-C36	22104953.000	180.000	13.051	18.471	122805.294	118321.901	-3.789
Aromatic EPH	47024670.000	360.000	4.388	18.471	130624.083	124871.589	-4.607