

Continuing Calibration Report for SequenceID : FF092525AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FF016451.D

Aromatic C10-C12	4713656.000	40.000	4.521	6.398	117841.400	143322.298	17.779
Aromatic C12-C16	7290904.000	60.000	6.399	9.102	121515.067	145185.730	16.304
Aromatic C16-C21	9645534.000	80.000	9.103	13.422	120569.175	144173.003	16.372
Aromatic C21-C36	19957430.000	180.000	13.423	18.876	110874.611	130834.072	15.256
Aromatic EPH	41607524.000	360.000	4.521	18.876	115576.456	137577.803	15.992