284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Alliance **Client:** G Environmental

Lab Code: ACE **SDG No:** Q3192

Sample No : PB169823BS **Datafile:** FC069855.D **Client ID :** PB169823BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS (%)
Aliphatic C9-C12	10	5.20	52		(40-140)
Aliphatic C12-C16	6.7	4.49	67		(40-140)
Aliphatic C16-C21	10	7.34	74		(40-140)
Aliphatic C21-C28	13.3	9.80	74		(40-140)
Aliphatic C28-C40	20.0	16.1	81		(40-140)