



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.:	Q3193	Analytical Method:	SW8270-Modified
Client:	AECOM Technical Services, Inc.	DataFile:	BN037948.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3193-02MS	Client Sample ID:	RW10-MW01D-20250923MS								
1,4-Dioxane	0.4	0	0.33	ug/L	83				70	130	