

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3193

Client ID : SSTDCCC0.4

Date Analyzed: 09/30/2025

Lab File ID: BN037943.D

Time Analyzed: 12:16

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	9644	7.811	27222	10.62	14738	14.46
UPPER LIMIT	19288	8.311	54444	11.116	29476	14.962
LOWER LIMIT	4822	7.311	13611	10.116	7369	13.962
EPA SAMPLE NO.						
01 PB169886BS	7452	7.81	21033	10.62	11010	14.46
02 RW10-MW01D-20250923	7881	7.81	23205	10.62	11948	14.47
03 RW10-MW01D-20250923MS	7993	7.81	22898	10.62	12112	14.47
04 RW10-MW01D-20250923MSD	6976	7.81	19964	10.62	10698	14.46
05 TT149S1-20250922	8551	7.81	24701	10.62	13182	14.46
06 RE137-20250922	7660	7.81	22207	10.62	11684	14.46
07 RW4-20250922	7793	7.81	22517	10.62	11801	14.46
08 GM38-RW3-MW1-20250922	7382	7.81	21403	10.62	11279	14.46
09 TT20551-20250923	8400	7.81	25968	10.62	13547	14.47
10 RW10A-MW01I-20250923	7049	7.81	21513	10.62	11227	14.46
11 RW8-MW01S-20250923	8340	7.81	24203	10.62	12799	14.46
12 TT172S1-20250923	7098	7.81	20498	10.62	10965	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.